


ISM02.4

LB89788
INSTRUMENT ID: CY1

Linear

 μ Abs.:

44204



A= 0.0000e+000

B= 2.2728e-004

C= -1.2096e-002

Rho= 0.9999330

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.023	-0.023	-50	0.000	-50					-
0.2	0.200	0.200	0.000	934	0.0 %	934					0
2.5	2.500	2.538	0.038	11222	0.0 %	11222					2
5.0	5.000	5.032	0.032	22193	0.0 %	22193					1
7.5	7.500	7.418	-0.082	32692	0.0 %	32692					-1
10.0	10.000	10.035	0.035	44204	0.0 %	44204					0

nb
8/30/13

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-50	-		0 ISM02.4	PPB	8/30/2017 12:46	Std
	0.2 S0.2	934	-		0.2 ISM02.4	PPB	8/30/2017 12:48	Std
	2.5 S2.5	11222	-		2.5 ISM02.4	PPB	8/30/2017 12:50	Std
	5 S5	22193	-		5 ISM02.4	PPB	8/30/2017 12:53	Std
	7.5 S7.5	32692	-		7.5 ISM02.4	PPB	8/30/2017 12:55	Std
	10 S10	44204	-		10 ISM02.4	PPB	8/30/2017 13:00	Std
ICV	ICV	17251	3.9087 -		ISM02.4	PPB	8/30/2017 13:02	SMPL
ICB	ICB	-221	-0.0623 -		ISM02.4	PPB	8/30/2017 13:04	SMPL
CCV089	CCV089	22269	5.0492 -		ISM02.4	PPB	8/30/2017 13:07	SMPL
CCB089	CCB089	-243	-0.0673 -		ISM02.4	PPB	8/30/2017 13:09	SMPL
PB101937BL	PBS004	-17	-0.016 -		ISM02.4	PPB	8/30/2017 13:11	SMPL
I5017-01	MET1T6	1575	0.3459 -		ISM02.4	PPB	8/30/2017 13:14	SMPL
I5017-02	MET1T7	6506	1.4666 -		ISM02.4	PPB	8/30/2017 13:16	SMPL
I5017-03	MET1T7D	5492	1.2361 -		ISM02.4	PPB	8/30/2017 13:18	SMPL
I5017-04	MET1T7S	17816	4.0372 -		ISM02.4	PPB	8/30/2017 13:20	SMPL
I5017-05	MET1T8	2014	0.4456 -		ISM02.4	PPB	8/30/2017 13:23	SMPL
I5017-06	MET1T9	3348	0.7488 -		ISM02.4	PPB	8/30/2017 13:25	SMPL
I5017-07	MET1W0	1495	0.3277 -		ISM02.4	PPB	8/30/2017 13:27	SMPL
I5017-08	MET1W1	3253	0.7273 -		ISM02.4	PPB	8/30/2017 13:29	SMPL
I5017-09	MET1W2	2188	0.4852 -		ISM02.4	PPB	8/30/2017 13:32	SMPL
I5017-10	MET1W3	3773	0.8454 -		ISM02.4	PPB	8/30/2017 13:34	SMPL
I5017-11	MET1W4	5013	1.1273 -		ISM02.4	PPB	8/30/2017 13:36	SMPL
I5017-12	MET1W5	1955	0.4322 -		ISM02.4	PPB	8/30/2017 13:39	SMPL
I5017-13	MET1W6	1943	0.4295 -		ISM02.4	PPB	8/30/2017 13:41	SMPL
I5017-14	MET1W7	3327	0.7441 -		ISM02.4	PPB	8/30/2017 13:43	SMPL
I5017-15	MET1W8	1871	0.4131 -		ISM02.4	PPB	8/30/2017 13:45	SMPL
I5017-16	MET1W9	1821	0.4018 -		ISM02.4	PPB	8/30/2017 13:48	SMPL
I5017-17	MET1X0	1976	0.437 -		ISM02.4	PPB	8/30/2017 13:50	SMPL
I5017-18	MET1X1	2823	0.6295 -		ISM02.4	PPB	8/30/2017 13:52	SMPL
I5017-19	MET1X2	1722	0.3793 -		ISM02.4	PPB	8/30/2017 13:54	SMPL
CCV090	CCV090	22782	5.1658 -		ISM02.4	PPB	8/30/2017 13:57	SMPL
CCB090	CCB090	-187	-0.0546 -		ISM02.4	PPB	8/30/2017 13:59	SMPL
I5017-20	MET1X3	1686	0.3711 -		ISM02.4	PPB	8/30/2017 14:01	SMPL
I5017-21	MET1X4	1654	0.3638 -		ISM02.4	PPB	8/30/2017 14:04	SMPL
CCV091	CCV091	22778	5.1649 -		ISM02.4	PPB	8/30/2017 14:06	SMPL
CCB091	CCB091	-204	-0.0585 -		ISM02.4	PPB	8/30/2017 14:08	SMPL