


7471A

LB89296
INSTRUMENT ID: CV4

Linear

 μ Abs.:

39957



A= 0.0000e+000

B= 2.5312e-004

C= 2.2637e-002

Rho= 0.9996416

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-97	0.000	-97				
0.2	0.200	0.230	0.030	820	0.0 %	820				
2.5	2.500	2.574	0.074	10081	0.0 %	10081				
5.0	5.000	4.929	-0.071	19383	0.0 %	19383				
7.5	7.500	7.332	-0.168	28875	0.0 %	28875	0			
10.0	10.000	10.137	0.137	39957	0.0 %	39957				

LB89796 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 30 Aug 2017 16:40:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-97	-	0.0000	7471A	PPB	30 Aug 2017 16:41:27	S	Std
0.2 - 1	50.2	820	-	0.2000	7471A	PPB	30 Aug 2017 16:43:46	S	Std
2.5 - 1	52.5	10081	-	2.5000	7471A	PPB	30 Aug 2017 16:46:04	S	Std
5.0 - 1	55	19383	-	5.0000	7471A	PPB	30 Aug 2017 16:48:21	S	Std
7.5 - 1	57.5	28875	-	7.5000	7471A	PPB	30 Aug 2017 16:50:38	S	Std
10.0 - 1	510	39957	-	10.0000	7471A	PPB	30 Aug 2017 16:55:30	S	Std
ICV36 - 1	ICV36	15815	4.0258	-7471A	PPB	30 Aug 2017 16:58:11	U	SMPL	
ICB36 - 1	ICB36	-252	-0.0412	-7471A	PPB	30 Aug 2017 17:00:26	U	SMPL	
CCV95 - 1	CCV95	19335	4.9168	-7471A	PPB	30 Aug 2017 17:02:43	U	SMPL	
CCB95 - 1	CCB95	-245	-0.0394	-7471A	PPB	30 Aug 2017 17:04:59	U	SMPL	
CRA - 1	CRA	790	0.2226	-7471A	PPB	30 Aug 2017 17:07:14	U	SMPL	
HighStd - 1	HighStd	40464	10.2651	-7471A	PPB	30 Aug 2017 17:09:30	U	SMPL	
ChkStd - 1	ChkStd	27428	6.9653	-7471A	PPB	30 Aug 2017 17:11:46	U	SMPL	
PB101932BL - 1	PBS	-279	-0.0480	-7471A	PPB	30 Aug 2017 17:14:05	U	SMPL	
PB101932BS - 1	LCSS	15142	3.8555	-7471A	PPB	30 Aug 2017 17:16:22	U	SMPL	
I5019-01 - 1	STAGE2-TEST-PIT-A-B-C	21143	5.3745	-7471A	PPB	30 Aug 2017 17:18:39	U	SMPL	
I5023-01 - 1	PL-01-082917-A	3713	0.9625	-7471A	PPB	30 Aug 2017 17:20:57	U	SMPL	
I5023-04 - 1	PL-01-082917-B	1147	0.3130	-7471A	PPB	30 Aug 2017 17:23:16	U	SMPL	
I5023-07 - 1	PL-01-082917-C	1634	0.4362	-7471A	PPB	30 Aug 2017 17:25:34	U	SMPL	
I5023-10 - 1	PL-01-082917-D	1591	0.4254	-7471A	PPB	30 Aug 2017 17:27:50	U	SMPL	
CCV96 - 1	CCV96	19584	4.9798	-7471A	PPB	30 Aug 2017 17:30:06	U	SMPL	
CCB96 - 1	CCB96	-218	-0.0325	-7471A	PPB	30 Aug 2017 17:32:22	U	SMPL	
I5023-10DUP - 1	PL-01-082917-DDUP	1815	0.4821	-7471A	PPB	30 Aug 2017 17:34:39	U	SMPL	
I5023-10MS - 1	PL-01-082917-DMS	17732	4.5110	-7471A	PPB	30 Aug 2017 17:36:55	U	SMPL	
I5023-10MSD - 1	PL-01-082917-DMSD	17645	4.4890	-7471A	PPB	30 Aug 2017 17:39:11	U	SMPL	
I5024-01 - 1	CL-01-082917-A	2431	0.6380	-7471A	PPB	30 Aug 2017 17:41:28	U	SMPL	
I5024-04 - 1	CL-01-082917-B	8325	2.1299	-7471A	PPB	30 Aug 2017 17:43:44	U	SMPL	
I5024-07 - 1	CL-01-082917-C	2931	0.7645	-7471A	PPB	30 Aug 2017 17:46:01	U	SMPL	
I5024-10 - 1	CL-01-082917-D	3645	0.9453	-7471A	PPB	30 Aug 2017 17:48:19	U	SMPL	
I5024-13 - 1	CL-01-082917-E	119072	30.1627	-7471A	PPB	30 Aug 2017 17:50:36	U	SMPL	
I5027-01 - 1	SU-02-082917	752	0.2130	-7471A	PPB	30 Aug 2017 17:52:54	U	SMPL	
CCV97 - 1	CCV97	19597	4.9831	-7471A	PPB	30 Aug 2017 17:55:12	U	SMPL	
CCB97 - 1	CCB97	-241	-0.0384	-7471A	PPB	30 Aug 2017 17:57:28	U	SMPL	
I5023-10LX5 - 1		171	0.0659	-7471A	PPB	30 Aug 2017 17:59:46	U	SMPL	
I5023-10A - 1		18314	4.6584	-7471A	PPB	30 Aug 2017 18:02:03	U	SMPL	
I5024-13DLX5 - 1		22456	5.7068	-7471A	PPB	30 Aug 2017 18:07:06	U	SMPL	
CCV98 - 1	CCV98	19274	4.9014	-7471A	PPB	30 Aug 2017 18:09:22	U	SMPL	
CCB98 - 1	CCB98	-285	-0.0495	-7471A	PPB	30 Aug 2017 18:11:39	U	SMPL	