


ISM02.4

LB89797
INSTRUMENT ID: CV2 μ Abs.:

41054



Linear

A= 0.0000e+000

B= 2.4486e-004

C= -1.0967e-002

Rho= 0.9999451

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.025	-0.025	-59	0.000	-59					1
0.2	0.200	0.214	0.014	920	0.0 %	920					7
2.5	2.500	2.543	0.043	10430	0.0 %	10430					2
5.0	5.000	4.994	-0.006	20439	0.0 %	20439					0
7.5	7.500	7.433	-0.067	30400	0.0 %	30400					-1
10.0	10.000	10.042	0.042	41054	0.0 %	41054					0

NB
8/30/12

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-59	-		0 ISM02.4	PPB	8/30/2017 18:28	Std
	0.2 S0.2	920	-		0.2 ISM02.4	PPB	8/30/2017 18:30	Std
	2.5 S2.5	10430	-		2.5 ISM02.4	PPB	8/30/2017 18:32	Std
	5 S5	20439	-		5 ISM02.4	PPB	8/30/2017 18:34	Std
	7.5 S7.5	30400	-		7.5 ISM02.4	PPB	8/30/2017 18:37	Std
	10 S10	41054	-		10 ISM02.4	PPB	8/30/2017 18:39	Std
ICV	ICV	15680	3.8285 -		ISM02.4	PPB	8/30/2017 18:42	SMPL
ICB	ICB	-241	-0.07 -		ISM02.4	PPB	8/30/2017 18:44	SMPL
CCV001	CCV001	20564	5.0244 -		ISM02.4	PPB	8/30/2017 18:46	SMPL
CCB001	CCB001	-212	-0.0629 -		ISM02.4	PPB	8/30/2017 18:48	SMPL
PB101952BL	PBW003	-239	-0.0695 -		ISM02.4	PPB	8/30/2017 18:51	SMPL
I5037-01	MBDCZ1	-294	-0.083 -		ISM02.4	PPB	8/30/2017 18:53	SMPL
I5037-02	MBDCZ4	-209	-0.0621 -		ISM02.4	PPB	8/30/2017 18:55	SMPL
I5037-03	MBDCZ4D	-230	-0.0673 -		ISM02.4	PPB	8/30/2017 18:58	SMPL
I5037-04	MBDCZ4S	4048	0.9802 -		ISM02.4	PPB	8/30/2017 19:00	SMPL
I5037-05	MBDCZ5	-257	-0.0739 -		ISM02.4	PPB	8/30/2017 19:02	SMPL
I5037-06	MBDCZ8	-295	-0.0832 -		ISM02.4	PPB	8/30/2017 19:04	SMPL
PB101953BL	PBW003	-173	-0.0533 -		ISM02.4	PPB	8/30/2017 19:07	SMPL
I5038-01	MBDCZ2	-351	-0.0969 -		ISM02.4	PPB	8/30/2017 19:09	SMPL
I5038-02	MBDCZ3	-311	-0.0871 -		ISM02.4	PPB	8/30/2017 19:11	SMPL
I5038-03	MBDCZ3D	-308	-0.0864 -		ISM02.4	PPB	8/30/2017 19:14	SMPL
I5038-04	MBDCZ3S	4026	0.9748 -		ISM02.4	PPB	8/30/2017 19:16	SMPL
I5038-05	MBDCZ6	-286	-0.081 -		ISM02.4	PPB	8/30/2017 19:18	SMPL
I5038-06	MBDCZ7	-277	-0.0788 -		ISM02.4	PPB	8/30/2017 19:20	SMPL
CCV002	CCV002	20423	4.9898 -		ISM02.4	PPB	8/30/2017 19:23	SMPL
CCB002	CCB002	-106	-0.0369 -		ISM02.4	PPB	8/30/2017 19:25	SMPL