

7470A


LB 89805

INSTRUMENT 1D: CV2

Linear

 μ Abs.:

36241



A= 0.0000e+000

B= 2.7660e-004

C= 4.8269e-002

Rho= 0.9999049

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-68	0.000	-68				
0.2	0.200	0.212	0.012	591	0.0 %	591				
2.5	2.500	2.517	0.017	8927	0.0 %	8927				
5.0	5.000	4.921	-0.079	17617	0.0 %	17617				
7.5	7.500	7.448	-0.052	26751	0.0 %	26751				
10.0	10.000	10.073	0.073	36241	0.0 %	36241				

LB89805 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Aug 2017 11:29:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.00007470A	PPB		31 Aug 2017 11:34:43	S	Std
0.2 - 1	50.2	591	-	0.20007470A	PPB		31 Aug 2017 11:36:51	S	Std
2.5 - 1	52.5	8927	-	2.50007470A	PPB		31 Aug 2017 11:38:59	S	Std
5.0 - 1	55	17617	-	5.00007470A	PPB		31 Aug 2017 11:41:08	S	Std
7.5 - 1	57.5	26751	-	7.50007470A	PPB		31 Aug 2017 11:43:16	S	Std
10.0 - 1	510	36241	-	10.00007470A	PPB		31 Aug 2017 11:45:23	S	Std
ICV37 - 1	ICV37	13703	3.8385	-7470A	PPB		31 Aug 2017 11:47:52	U	SMPL
ICB37 - 1	ICB37	-182	-0.0021	-7470A	PPB		31 Aug 2017 11:49:58	U	SMPL
CCV03 - 1	CCV03	17771	4.9637	-7470A	PPB		31 Aug 2017 11:52:05	U	SMPL
CCB03 - 1	CCB03	-184	-0.0026	-7470A	PPB		31 Aug 2017 11:54:11	U	SMPL
CRA - 1	CRA	549	0.2001	-7470A	PPB		31 Aug 2017 11:56:19	U	SMPL
HighStd - 1	HighStd	37171	10.3298	-7470A	PPB		31 Aug 2017 11:58:25	U	SMPL
ChkStd - 1	ChkStd	25842	7.1962	-7470A	PPB		31 Aug 2017 12:00:32	U	SMPL
PB101954BL - 1	PBW	-234	-0.0165	-7470A	PPB		31 Aug 2017 12:02:41	U	SMPL
PB101954BS - 1	LCWS	13902	3.8936	-7470A	PPB		31 Aug 2017 12:04:49	U	SMPL
PB101954TB - 1	PB101926TB	-153	0.0059	-7470A	PPB		31 Aug 2017 12:06:57	U	SMPL
I5019-07 - 1	STAGE2-TEST-PIT-A-B-C	-136	0.0107	-7470A	PPB		31 Aug 2017 12:09:05	U	SMPL
I5023-03 - 1	PL-01-082917-A	-134	0.0112	-7470A	PPB		31 Aug 2017 12:11:13	U	SMPL
I5023-06 - 1	PL-01-082917-B	-105	0.0192	-7470A	PPB		31 Aug 2017 12:13:22	U	SMPL
I5023-09 - 1	PL-01-082917-C	-108	0.0184	-7470A	PPB		31 Aug 2017 12:15:29	U	SMPL
CCV04 - 1	CCV04	18052	5.0415	-7470A	PPB		31 Aug 2017 12:17:35	U	SMPL
CCB04 - 1	CCB04	-155	0.0054	-7470A	PPB		31 Aug 2017 12:19:42	U	SMPL
I5023-12 - 1	PL-01-082917-D	-107	0.0187	-7470A	PPB		31 Aug 2017 12:21:49	U	SMPL
I5024-03 - 1	CL-01-082917-A	-92	0.0228	-7470A	PPB		31 Aug 2017 12:23:56	U	SMPL
I5024-06 - 1	CL-01-082917-B	-75	0.0275	-7470A	PPB		31 Aug 2017 12:26:03	U	SMPL
I5024-09 - 1	CL-01-082917-C	-88	0.0239	-7470A	PPB		31 Aug 2017 12:28:10	U	SMPL
I5024-12 - 1	CL-01-082917-D	-70	0.0289	-7470A	PPB		31 Aug 2017 12:30:17	U	SMPL
I5024-15 - 1	CL-01-082917-E	-51	0.0342	-7470A	PPB		31 Aug 2017 12:32:25	U	SMPL
I5027-03 - 1	SU-02-082917	-67	0.0297	-7470A	PPB		31 Aug 2017 12:34:33	U	SMPL
I5027-03DUP - 1	SU-02-082917DUP	-88	0.0239	-7470A	PPB		31 Aug 2017 12:36:42	U	SMPL
I5027-03MS - 1	SU-02-082917MS	12884	3.6120	-7470A	PPB		31 Aug 2017 12:38:50	U	SMPL
I5027-03MSD - 1	SU-02-082917MSD	12877	3.6101	-7470A	PPB		31 Aug 2017 12:40:57	U	SMPL
CCV05 - 1	CCV05	18151	5.0688	-7470A	PPB		31 Aug 2017 12:43:04	U	SMPL
CCB05 - 1	CCB05	-191	-0.0046	-7470A	PPB		31 Aug 2017 12:45:11	U	SMPL
I5027-03LX5 - 1		-107	0.0187	-7470A	PPB		31 Aug 2017 12:47:18	U	SMPL
I5027-03A - 1		14249	3.9896	-7470A	PPB		31 Aug 2017 12:49:24	U	SMPL
CCV06 - 1	CCV06	17915	5.0036	-7470A	PPB		31 Aug 2017 12:51:31	U	SMPL
CCB06 - 1	CCB06	-182	-0.0021	-7470A	PPB		31 Aug 2017 12:53:38	U	SMPL