


7471A

LB89812
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43407



A= 0.0000e+000

B= 2.3374e-004

C= -1.9233e-002

Rho= 0.9996912

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	49	0.000	49				
0.2	0.200	0.239	0.039	1106	0.0 %	1106				
2.5	2.500	2.538	0.038	10941	0.0 %	10941				
5.0	5.000	4.979	-0.021	21385	0.0 %	21385				
7.5	7.500	7.324	-0.176	31417	0.0 %	31417	0			
10.0	10.000	10.127	0.127	43407	0.0 %	43407				

LB89812 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 31 Aug 2017 15:13:14

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0007471A	PPB	31 Aug 2017 15:21:08	S	Std
0.2 - 1	50.2	1106	-	0.20007471A	PPB	31 Aug 2017 15:23:26	S	Std
2.5 - 1	52.5	10941	-	2.50007471A	PPB	31 Aug 2017 15:25:43	S	Std
5.0 - 1	55	21385	-	5.00007471A	PPB	31 Aug 2017 15:28:01	S	Std
7.5 - 1	57.5	31417	-	7.50007471A	PPB	31 Aug 2017 15:30:18	S	Std
10.0 - 1	60	43407	-	10.00007471A	PPB	31 Aug 2017 15:35:16	S	Std
ICV38 - 1	ICV38	16524	3.8431	-7471A	PPB	31 Aug 2017 15:38:07	U	SMPL
ICB38 - 1	ICB38	-46	-0.0300	-7471A	PPB	31 Aug 2017 15:40:23	U	SMPL
CCV07 - 1	CCV07	21215	4.9396	-7471A	PPB	31 Aug 2017 15:42:39	U	SMPL
CCB07 - 1	CCB07	-36	-0.0276	-7471A	PPB	31 Aug 2017 15:44:54	U	SMPL
CRA - 1	CRA	1055	0.2274	-7471A	PPB	31 Aug 2017 15:47:10	U	SMPL
HighStd - 1	HighStd	43530	10.1555	-7471A	PPB	31 Aug 2017 15:49:26	U	SMPL
ChkStd - 1	ChkStd	30580	7.1286	-7471A	PPB	31 Aug 2017 15:51:42	U	SMPL
PB101958BL - 1	PBS	-202	-0.0664	-7471A	PPB	31 Aug 2017 15:54:00	U	SMPL
PB101958BS - 1	LCSS	15322	3.5622	-7471A	PPB	31 Aug 2017 15:56:18	U	SMPL
I5025-05 - 1	TEC-COMP-3-6	346	0.0616	-7471A	PPB	31 Aug 2017 15:58:35	U	SMPL
I5025-05DUP - 1	TEC-COMP-3-6DUP	267	0.0432	-7471A	PPB	31 Aug 2017 16:00:52	U	SMPL
I5025-05MS - 1	TEC-COMP-3-6MS	14735	3.4250	-7471A	PPB	31 Aug 2017 16:03:10	U	SMPL
I5025-05MSD - 1	TEC-COMP-3-6MSD	14883	3.4595	-7471A	PPB	31 Aug 2017 16:05:28	U	SMPL
I5025-06 - 1	TEC-COMP-0-3	2646	0.5992	-7471A	PPB	31 Aug 2017 16:07:44	U	SMPL
CCV08 - 1	CCV08	21184	4.9324	-7471A	PPB	31 Aug 2017 16:10:00	U	SMPL
CCB08 - 1	CCB08	-13	-0.0223	-7471A	PPB	31 Aug 2017 16:12:16	U	SMPL
I5025-07 - 1	TEC-COMP-6-9	998	0.2140	-7471A	PPB	31 Aug 2017 16:14:32	U	SMPL
I5025-08 - 1	TEC-COMP-9-12	433	0.0820	-7471A	PPB	31 Aug 2017 16:16:48	U	SMPL
I5026-02 - 1	B-17(0-10)COMP	9769	2.2642	-7471A	PPB	31 Aug 2017 16:19:04	U	SMPL
I5028-01 - 1	EP-04	59	-0.0054	-7471A	PPB	31 Aug 2017 16:21:20	U	SMPL
I5046-02 - 1	ROLLOFF-342	289	0.0483	-7471A	PPB	31 Aug 2017 16:23:37	U	SMPL
I5047-02 - 1	LEAD-PAINT-CHIPS-DEBRIS-DRUM	2854	0.6479	-7471A	PPB	31 Aug 2017 16:25:54	U	SMPL
I5048-02 - 1	OILY-RAGS-DEBRIS	227	0.0338	-7471A	PPB	31 Aug 2017 16:28:11	U	SMPL
I5048-04 - 1	OILY-STONES	-8	-0.0211	-7471A	PPB	31 Aug 2017 16:30:28	U	SMPL
I5025-05LX5 - 1		172	0.0210	-7471A	PPB	31 Aug 2017 16:32:46	U	SMPL
I5025-05A - 1		17257	4.0144	-7471A	PPB	31 Aug 2017 16:35:02	U	SMPL
CCV09 - 1	CCV09	21124	4.9183	-7471A	PPB	31 Aug 2017 16:37:18	U	SMPL
CCB09 - 1	CCB09	-19	-0.0237	-7471A	PPB	31 Aug 2017 16:39:34	U	SMPL