



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89821

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/31/2017	17:25
2	CAL2	1	Water			7.01	08/31/2017	17:26
3	CAL3	1	Water			10.04	08/31/2017	17:27
4	ICV	1	Water			7.00	08/31/2017	17:30
5	CCV1	1	Water			2.01	08/31/2017	17:31
6	I5036-03	1	Solid	20.02	20	5.95	08/31/2017	17:33
7	I5049-03	1	Solid	20.04	20	7.21	08/31/2017	17:35
8	I5050-04	1	Solid	20.02	20	7.23	08/31/2017	17:36
9	I5051-03	1	Solid	20.05	20	9.20	08/31/2017	17:37
10	I5055-06	1	Solid	20.03	20	7.87	08/31/2017	17:40
11	I5068-03	1	Solid	20.01	20	9.04	08/31/2017	17:41
12	I5069-03	1	Solid	20.02	20	9.14	08/31/2017	17:42
13	I5073-02	1	Solid	20.03	20	7.78	08/31/2017	17:44
14	I5073-09	1	Solid	20.02	20	8.08	08/31/2017	17:46
15	I5074-02	1	Solid	20.03	20	7.69	08/31/2017	17:48
16	CCV2	1	Water			12.02	08/31/2017	17:49
17	I5074-09	1	Solid	20.02	20	7.58	08/31/2017	17:50
18	I5074-16	1	Solid	20.02	20	7.37	08/31/2017	17:53
19	I5074-16DUP	1	Solid	20.03	20	7.38	08/31/2017	17:55
20	CCV3	1	Water			12.01	08/31/2017	18:00

13 89821

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5055

WorkList ID : 102459

Date : 8/31/2017 12:57:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2017	Solid	I5036-03	Corrosivity	Cool 4 deg C	PSEG01	O53	T2-T3-COMP	08/30/2017	9045C
	Solid	I5049-03	Corrosivity	Cool 4 deg C	PSEG05	O53	TR-04-083017	08/30/2017	9045C
	Solid	I5050-04	Corrosivity	Cool 4 deg C	PSEG01	O52	RT4651-1540	08/30/2017	9045C
	Solid	I5051-03	Corrosivity	Cool 4 deg C	TECT01	O41	TEC-B2-COMP	08/30/2017	9045C
	Solid	I5055-06	Corrosivity	Cool 4 deg C	PSEG01	O52	RT3992-20-41-RT3451-1550	08/30/2017	9045C
	Solid	I5068-03	Corrosivity	Cool 4 deg C	PSEG05	O53	HR-06-083117	08/31/2017	9045C
	Solid	I5069-03	Corrosivity	Cool 4 deg C	DVIR01	O52	I115	08/31/2017	9045C
	Solid	I5073-02	Corrosivity	Cool 4 deg C	PSEG01	O53	TP-1-2-COMP	08/31/2017	9045C
	Solid	I5073-09	Corrosivity	Cool 4 deg C	PSEG01	O53	TP-3-4-COMP	08/31/2017	9045C
	Solid	I5074-02	Corrosivity	Cool 4 deg C	PSEG01	O53	TP-5-6-COMP	08/31/2017	9045C
	Solid	I5074-09	Corrosivity	Cool 4 deg C	PSEG01	O53	TP-7-8-COMP	08/31/2017	9045C
	Solid	I5074-16	Corrosivity	Cool 4 deg C	PSEG01	O53	TP-9-10-COMP	08/31/2017	9045C

Date/Time 08/31/17 1:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 08/31/17 7:00 PM
 Received by: [Signature]
 Relinquished by: [Signature]