


ISM02.4

LB89826
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4573E



A= 0.0000e+000

B= 2.1890e-004

C= -3.1483e-002

Rho= 0.9999626

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.040	-0.040	-40	0.000	-40					
0.2	0.200	0.190	-0.010	1011	0.0 %	1011					
2.5	2.500	2.563	0.063	11851	0.0 %	11851					
5.0	5.000	5.008	0.008	23020	0.0 %	23020					
7.5	7.500	7.500	-0.000	34406	0.0 %	34406					
10.0	10.000	9.980	-0.020	45736	0.0 %	45736					

LB89826
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-40	-		0 ISM02.4	PPB	9/1/2017 10:03	Std
	0.2 S0.2	1011	-		0.2 ISM02.4	PPB	9/1/2017 10:05	Std
	2.5 S2.5	11851	-		2.5 ISM02.4	PPB	9/1/2017 10:08	Std
	5 S5	23020	-		5 ISM02.4	PPB	9/1/2017 10:10	Std
	7.5 S7.5	34406	-		7.5 ISM02.4	PPB	9/1/2017 10:12	Std
	10 S10	45736	-		10 ISM02.4	PPB	9/1/2017 10:14	Std
ICV	ICV	18002	3.9092 -		ISM02.4	PPB	9/1/2017 10:17	SMPL
ICB	ICB	-223	-0.0803 -		ISM02.4	PPB	9/1/2017 10:20	SMPL
CCV012	CCV012	23122	5.0299 -		ISM02.4	PPB	9/1/2017 10:22	SMPL
CCB012	CCB012	-221	-0.0799 -		ISM02.4	PPB	9/1/2017 10:24	SMPL
PB101992BL	PBS004	42	-0.0223 -		ISM02.4	PPB	9/1/2017 10:26	SMPL
I5061-01	MET1X5	7507	1.6118 -		ISM02.4	PPB	9/1/2017 10:29	SMPL
I5061-02	MET1X6	4711	0.9998 -		ISM02.4	PPB	9/1/2017 10:31	SMPL
I5061-03	MET1X7	2902	0.6038 -		ISM02.4	PPB	9/1/2017 10:33	SMPL
I5061-04	MET1X8	2574	0.532 -		ISM02.4	PPB	9/1/2017 10:35	SMPL
I5061-05	MET1X9	2823	0.5865 -		ISM02.4	PPB	9/1/2017 10:38	SMPL
I5061-06	MET1Y0	3847	0.8106 -		ISM02.4	PPB	9/1/2017 10:40	SMPL
I5061-07	MET1Y1	4589	0.973 -		ISM02.4	PPB	9/1/2017 10:42	SMPL
I5061-08	MET1Y1D	4579	0.9709 -		ISM02.4	PPB	9/1/2017 10:45	SMPL
I5061-09	MET1Y1S	16218	3.5186 -		ISM02.4	PPB	9/1/2017 10:47	SMPL
I5061-10	MET1Y2	2208	0.4518 -		ISM02.4	PPB	9/1/2017 10:49	SMPL
I5061-11	MET1Y3	2800	0.5814 -		ISM02.4	PPB	9/1/2017 10:51	SMPL
I5061-12	MET1Y4	1806	0.3639 -		ISM02.4	PPB	9/1/2017 10:54	SMPL
I5061-13	MET1Y5	2297	0.4713 -		ISM02.4	PPB	9/1/2017 10:56	SMPL
I5061-14	MET1Y6	2008	0.4081 -		ISM02.4	PPB	9/1/2017 10:58	SMPL
I5061-15	MET1Y7	2690	0.5574 -		ISM02.4	PPB	9/1/2017 11:00	SMPL
I5061-16	MET1Y8	3317	0.6946 -		ISM02.4	PPB	9/1/2017 11:03	SMPL
I5061-17	MET1Y9	7660	1.6453 -		ISM02.4	PPB	9/1/2017 11:05	SMPL
I5061-18	MET1Z0	8156	1.7539 -		ISM02.4	PPB	9/1/2017 11:07	SMPL
CCV013	CCV013	23322	5.0737 -		ISM02.4	PPB	9/1/2017 11:10	SMPL
CCB013	CCB013	-242	-0.0845 -		ISM02.4	PPB	9/1/2017 11:12	SMPL