

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB89830

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP57628
calibration std. phosphate 0.5 ppm	WP57627
calibration std. phosphate 0.3 ppm	WP57626
calibration std. 0 ppm	WP57623
phosphate ICV std.	WP57630
calibration std. phosphate 0.1 ppm	WP57625
calibration std. phosphate 0.05 ppm	WP57624
Combined reagent	WP57632
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036
phosphate CCV std.	WP57629
Ascorbic Acid	WP57631

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	09/01/2017	11:20
2	CAL2	0.05	1	50	50	0.033	09/01/2017	11:20
3	CAL3	0.10	1	50	50	0.065	09/01/2017	11:21
4	CAL4	0.30	1	50	50	0.186	09/01/2017	11:21
5	CAL5	0.50	1	50	50	0.330	09/01/2017	11:22
6	CAL6	1.00	1	50	50	0.649	09/01/2017	11:22



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB89830

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.503	09/01/2017	11:23
2	ICB		1	50	50	0.000	0.001	09/01/2017	11:23
3	CCV1	0.50	1	50	50	0.329	0.507	09/01/2017	11:24
4	CCB1		1	50	50	0.000	0.001	09/01/2017	11:24
5	PB102008BL		1	50	50	0.000	0.001	09/01/2017	11:25
6	PB102008BS	0.50	1	50	50	0.328	0.506	09/01/2017	11:25
7	I5031-01		1	50	50	0.006	0.011	09/01/2017	11:26
8	I5031-01DUP		1	50	50	0.005	0.009	09/01/2017	11:26
9	I5031-01MS	0.50	1	50	50	0.347	0.535	09/01/2017	11:27
10	I5031-01MSD	0.50	1	50	50	0.345	0.532	09/01/2017	11:27
11	I5031-02		1	50	50	0.008	0.014	09/01/2017	11:28
12	CCV2	0.50	1	50	50	0.328	0.506	09/01/2017	11:28
13	CCB2		1	50	50	0.000	0.001	09/01/2017	11:29