


7470A

LB89891
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35356



A= 0.0000e+000

B= 2.8024e-004

C= 2.2762e-002

Rho= 0.9995806

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	-126	0.000	-126				
0.2	0.200	0.221	0.021	707	0.0 %	707				
2.5	2.500	2.552	0.052	9026	0.0 %	9026				
5.0	5.000	4.831	-0.169	17156	0.0 %	17156				
7.5	7.500	7.678	0.178	27317	0.0 %	27317				
10.0	10.000	9.931	-0.069	35356	0.0 %	35356				

LB89831 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Sep 2017 12:15:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	52	-126	-	0.00007470A	PPB		01 Sep 2017 12:16:54	S	Std
0.2 - 1	50.2	707	-	0.20007470A	PPB		01 Sep 2017 12:19:03	S	Std
2.5 - 1	52.5	9026	-	2.50007470A	PPB		01 Sep 2017 12:21:11	S	Std
5.0 - 1	55	17156	-	5.00007470A	PPB		01 Sep 2017 12:23:20	S	Std
7.5 - 1	57.5	27317	-	7.50007470A	PPB		01 Sep 2017 12:25:28	S	Std
10.0 - 1	610	35356	-	10.00007470A	PPB		01 Sep 2017 12:27:35	S	Std
ICV39 - 1	ICV39	13581	3.8287	-7470A	PPB		01 Sep 2017 12:30:09	U	SMPL
ICB39 - 1	ICB39	-205	-0.0347	-7470A	PPB		01 Sep 2017 12:32:15	U	SMPL
CCV14 - 1	CCV14	17149	4.8286	-7470A	PPB		01 Sep 2017 12:34:22	U	SMPL
CCB14 - 1	CCB14	-235	-0.0431	-7470A	PPB		01 Sep 2017 12:36:28	U	SMPL
CRA - 1	CRA	661	0.2080	-7470A	PPB		01 Sep 2017 12:38:35	U	SMPL
HighStd - 1	HighStd	35732	10.0363	-7470A	PPB		01 Sep 2017 12:40:41	U	SMPL
ChkStd - 1	ChkStd	24448	6.8740	-7470A	PPB		01 Sep 2017 12:42:48	U	SMPL
PB101985BL - 1	PBW	-281	-0.0560	-7470A	PPB		01 Sep 2017 12:44:57	U	SMPL
PB101985BS - 1	LCSW	13438	3.7886	-7470A	PPB		01 Sep 2017 12:47:05	U	SMPL
PB101985TB - 1	PB101948TB	-203	-0.0341	-7470A	PPB		01 Sep 2017 12:49:13	U	SMPL
I5026-03 - 1	B-17(0-10)COMP	-182	-0.0282	-7470A	PPB		01 Sep 2017 12:51:21	U	SMPL
I5026-03DUP - 1	B-17(0-10)COMP DUP	-146	-0.0182	-7470A	PPB		01 Sep 2017 12:53:30	U	SMPL
I5026-03MS - 1	B-17(0-10)COMPMS	13073	3.6863	-7470A	PPB		01 Sep 2017 12:55:38	U	SMPL
I5026-03MSD - 1	B-17(0-10)COMPMSD	12822	3.6160	-7470A	PPB		01 Sep 2017 12:57:45	U	SMPL
CCV15 - 1	CCV15	17007	4.7888	-7470A	PPB		01 Sep 2017 12:59:51	U	SMPL
CCB15 - 1	CCB15	-216	-0.0378	-7470A	PPB		01 Sep 2017 13:01:58	U	SMPL
I5046-01 - 1	ROLLOFF-342	-136	-0.0154	-7470A	PPB		01 Sep 2017 13:04:05	U	SMPL
I5047-01 - 1	LEAD-PAINT-CHIPS-DEBRIS-DRUM	-90	-0.0025	-7470A	PPB		01 Sep 2017 13:06:12	U	SMPL
I5048-01 - 1	OILY-RAGS-DEBRIS	-125	-0.0123	-7470A	PPB		01 Sep 2017 13:08:19	U	SMPL
I5048-03 - 1	OILY-STONES	-120	-0.0109	-7470A	PPB		01 Sep 2017 13:10:26	U	SMPL
I5026-03LX5 - 1		-128	-0.0131	-7470A	PPB		01 Sep 2017 13:12:33	U	SMPL
I5026-03A - 1		13869	3.9094	-7470A	PPB		01 Sep 2017 13:14:41	U	SMPL
CCV16 - 1	CCV16	17098	4.8143	-7470A	PPB		01 Sep 2017 13:16:50	U	SMPL
CCB16 - 1	CCB16	-231	-0.0420	-7470A	PPB		01 Sep 2017 13:18:58	U	SMPL