

7471A


LB 89836

INSTRUMENT ID: CV2

Linear

 μ Abs.:

43139



A= 0.0000e+000

B= 2.3368e-004

C= 1.2659e-002

Rho= 0.9998412

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	47	0.000	47				
0.2	0.200	0.222	0.022	894	0.0 %	894				
2.5	2.500	2.533	0.033	10787	0.0 %	10787				
5.0	5.000	4.903	-0.097	20928	0.0 %	20928				
7.5	7.500	7.425	-0.075	31718	0.0 %	31718	0			
10.0	10.000	10.094	0.094	43139	0.0 %	43139				

LB89836 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 01 Sep 2017 14:05:32

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471A	PPB		01 Sep 2017 14:07:30	S	Std
0.2 - 1	50.2	894	-	0.20007471A	PPB		01 Sep 2017 14:09:47	S	Std
2.5 - 1	52.5	10787	-	2.50007471A	PPB		01 Sep 2017 14:12:05	S	Std
5.0 - 1	55	20928	-	5.00007471A	PPB		01 Sep 2017 14:14:22	S	Std
7.5 - 1	57.5	31718	-	7.50007471A	PPB		01 Sep 2017 14:16:40	S	Std
10.0 - 1	50	43139	-	10.00007471A	PPB		01 Sep 2017 14:21:30	S	Std
ICV40 - 1	ICV40	16178	3.7932	-7471A	PPB		01 Sep 2017 14:24:09	U	SMPL
ICB40 - 1	ICB40	-143	-0.0208	-7471A	PPB		01 Sep 2017 14:26:25	U	SMPL
CCV17 - 1	CCV17	21880	5.1257	-7471A	PPB		01 Sep 2017 14:28:41	U	SMPL
CCB17 - 1	CCB17	-188	-0.0313	-7471A	PPB		01 Sep 2017 14:30:56	U	SMPL
CRA - 1	CRA	816	0.2033	-7471A	PPB		01 Sep 2017 14:33:12	U	SMPL
HighStd - 1	HighStd	43565	10.1931	-7471A	PPB		01 Sep 2017 14:35:28	U	SMPL
ChkStd - 1	ChkStd	30382	7.1124	-7471A	PPB		01 Sep 2017 14:37:43	U	SMPL
PB101993BL - 1	PBS	-252	-0.0462	-7471A	PPB		01 Sep 2017 14:40:02	U	SMPL
PB101993BS - 1	LCSS	15063	3.5326	-7471A	PPB		01 Sep 2017 14:42:20	U	SMPL
I5036-01 - 1	T2-T3-COMP	-109	-0.0128	-7471A	PPB		01 Sep 2017 14:44:37	U	SMPL
I5049-01 - 1	TR-04-083017	2127	0.5097	-7471A	PPB		01 Sep 2017 14:46:54	U	SMPL
I5050-01 - 1	RT4651-1540	2350	0.5618	-7471A	PPB		01 Sep 2017 14:49:12	U	SMPL
I5051-01 - 1	TEC-B2-COMP	19180	4.4947	-7471A	PPB		01 Sep 2017 14:51:29	U	SMPL
I5055-01 - 1	RT3992-20-41-RT3451-1550	1842	0.4431	-7471A	PPB		01 Sep 2017 14:53:45	U	SMPL
CCV18 - 1	CCV18	21329	4.9969	-7471A	PPB		01 Sep 2017 14:56:01	U	SMPL
CCB18 - 1	CCB18	-112	-0.0135	-7471A	PPB		01 Sep 2017 14:58:17	U	SMPL
I5055-01DUP - 1	RT3992-20-41-RT3451-1550DUP	1977	0.4747	-7471A	PPB		01 Sep 2017 15:03:17	U	SMPL
I5055-01MS - 1	RT3992-20-41-RT3451-1550MS	17309	4.0575	-7471A	PPB		01 Sep 2017 15:05:32	U	SMPL
I5055-01MSD - 1	RT3992-20-41-RT3451-1550MSD	17312	4.0582	-7471A	PPB		01 Sep 2017 15:07:48	U	SMPL
I5066-01 - 1	CF-1	698	0.1758	-7471A	PPB		01 Sep 2017 15:10:05	U	SMPL
I5068-01 - 1	HR-06-083117	13494	3.1660	-7471A	PPB		01 Sep 2017 15:12:21	U	SMPL
I5069-01 - 1	I115	2710	0.6459	-7471A	PPB		01 Sep 2017 15:14:38	U	SMPL
I5073-01 - 1	TP-1-2-COMP	5951	1.4033	-7471A	PPB		01 Sep 2017 15:16:55	U	SMPL
I5073-08 - 1	TP-3-4-COMP	5278	1.2460	-7471A	PPB		01 Sep 2017 15:19:12	U	SMPL
I5074-01 - 1	TP-5-6-COMP	17595	4.1243	-7471A	PPB		01 Sep 2017 15:21:30	U	SMPL
I5074-08 - 1	TP-7-8-COMP	5164	1.2194	-7471A	PPB		01 Sep 2017 15:23:46	U	SMPL
CCV19 - 1	CCV19	21312	4.9929	-7471A	PPB		01 Sep 2017 15:26:02	U	SMPL
CCB19 - 1	CCB19	-148	-0.0219	-7471A	PPB		01 Sep 2017 15:28:19	U	SMPL
I5074-15 - 1	TP-9-10-COMP	2456	0.5866	-7471A	PPB		01 Sep 2017 15:30:36	U	SMPL
I5055-01LX5 - 1		360	0.0968	-7471A	PPB		01 Sep 2017 15:32:52	U	SMPL
I5055-01A - 1		19409	4.5482	-7471A	PPB		01 Sep 2017 15:35:09	U	SMPL
CCV20 - 1	CCV20	21277	4.9847	-7471A	PPB		01 Sep 2017 15:37:25	U	SMPL
CCB20 - 1	CCB20	-152	-0.0229	-7471A	PPB		01 Sep 2017 15:39:42	U	SMPL