

LB 89840


ISM02.4

INSTRUMENT ID: CV4

Linear

 μ Abs.:

41380



A= 0.0000e+000

B= 2.4177e-004

C= 5.1334e-003

Rho= 0.9999916

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.003	-0.003	-34	0.000	-34					1
0.2	0.200	0.215	0.015	867	0.0 %	867					7
2.5	2.500	2.479	-0.021	10234	0.0 %	10234					-1
5.0	5.000	5.018	0.018	20732	0.0 %	20732					0
7.5	7.500	7.482	-0.018	30925	0.0 %	30925					0
10.0	10.000	10.010	0.010	41380	0.0 %	41380					0

MSB
9/11/12

LB89840
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-34	-		0 ISM02.4	PPB	9/1/2017 15:53	Std
	0.2 S0.2	867	-		0.2 ISM02.4	PPB	9/1/2017 15:56	Std
	2.5 S2.5	10234	-		2.5 ISM02.4	PPB	9/1/2017 16:00	Std
	5 S5	20732	-		5 ISM02.4	PPB	9/1/2017 16:02	Std
	7.5 S7.5	30925	-		7.5 ISM02.4	PPB	9/1/2017 16:04	Std
	10 S10	41380	-		10 ISM02.4	PPB	9/1/2017 16:07	Std
ICV	ICV	15981	3.8689 -		ISM02.4	PPB	9/1/2017 16:09	SMPL
ICB	ICB	-195	-0.042 -		ISM02.4	PPB	9/1/2017 16:12	SMPL
CCV021	CCV021	20385	4.9336 -		ISM02.4	PPB	9/1/2017 16:14	SMPL
CCB021	CCB021	-187	-0.0401 -		ISM02.4	PPB	9/1/2017 16:16	SMPL
PB102009BL	PBW010	-5	0.0039 -		ISM02.4	PPB	9/1/2017 16:18	SMPL
PB102009TB	LEB010	420	0.1067 -		ISM02.4	PPB	9/1/2017 16:21	SMPL
I5039-01	MA41D0	1748	0.4277 -		ISM02.4	PPB	9/1/2017 16:23	SMPL
I5039-02	MA41D0D	1670	0.4089 -		ISM02.4	PPB	9/1/2017 16:25	SMPL
I5039-03	MA41D0S	5688	1.3803 -		ISM02.4	PPB	9/1/2017 16:27	SMPL
I5039-04	MA41D1	8782	2.1284 -		ISM02.4	PPB	9/1/2017 16:30	SMPL
I5039-05	MA41D2	13714	3.3208 -		ISM02.4	PPB	9/1/2017 16:32	SMPL
I5039-06	MA41D3	3410	0.8296 -		ISM02.4	PPB	9/1/2017 16:34	SMPL
I5039-07	MA41D4	5072	1.2314 -		ISM02.4	PPB	9/1/2017 16:37	SMPL
I5039-08	MA41D5	1439	0.353 -		ISM02.4	PPB	9/1/2017 16:39	SMPL
I5039-09	MA41D6	955	0.236 -		ISM02.4	PPB	9/1/2017 16:41	SMPL
I5039-10	MA41D7	226	0.0598 -		ISM02.4	PPB	9/1/2017 16:43	SMPL
I5039-11	MA41D8	862	0.2135 -		ISM02.4	PPB	9/1/2017 16:46	SMPL
I5039-12	MA41D9	156	0.0428 -		ISM02.4	PPB	9/1/2017 16:48	SMPL
I5039-13	MA41F0	1173	0.2887 -		ISM02.4	PPB	9/1/2017 16:50	SMPL
I5039-14	MA41F1	524	0.1318 -		ISM02.4	PPB	9/1/2017 16:53	SMPL
I5039-15	MA41F2	2824	0.6879 -		ISM02.4	PPB	9/1/2017 16:55	SMPL
I5039-16	MA41F3	248	0.0651 -		ISM02.4	PPB	9/1/2017 16:57	SMPL
I5039-17	MA41F4	420	0.1067 -		ISM02.4	PPB	9/1/2017 16:59	SMPL
I5039-18	MA41F5	1448	0.3552 -		ISM02.4	PPB	9/1/2017 17:02	SMPL
CCV022	CCV022	20404	4.9382 -		ISM02.4	PPB	9/1/2017 17:04	SMPL
CCB022	CCB022	-189	-0.0406 -		ISM02.4	PPB	9/1/2017 17:06	SMPL
I5039-19	MA41F6	673	0.1678 -		ISM02.4	PPB	9/1/2017 17:09	SMPL
I5039-20	MA41F7	2426	0.5917 -		ISM02.4	PPB	9/1/2017 17:11	SMPL
I5039-21	MA41F8	148	0.0409 -		ISM02.4	PPB	9/1/2017 17:13	SMPL
I5039-22	MA41F9	2473	0.603 -		ISM02.4	PPB	9/1/2017 17:15	SMPL
PB102010BL	PBW010	-86	-0.0157 -		ISM02.4	PPB	9/1/2017 17:18	SMPL
PB102010TB	LEB010	-12	0.0022 -		ISM02.4	PPB	9/1/2017 17:20	SMPL
I5040-01	MA41E0	806	0.2 -		ISM02.4	PPB	9/1/2017 17:22	SMPL
I5040-02	MA41E1	299	0.0774 -		ISM02.4	PPB	9/1/2017 17:24	SMPL
I5040-03	MA41E2	-171	-0.0362 -		ISM02.4	PPB	9/1/2017 17:27	SMPL
I5040-04	MA41E3	94	0.0279 -		ISM02.4	PPB	9/1/2017 17:29	SMPL
I5040-05	MA41E3D	126	0.0356 -		ISM02.4	PPB	9/1/2017 17:31	SMPL
I5040-06	MA41E3S	4421	1.074 -		ISM02.4	PPB	9/1/2017 17:34	SMPL
I5040-07	MA41E4	-101	-0.0193 -		ISM02.4	PPB	9/1/2017 17:36	SMPL
I5040-08	MA41E5	-184	-0.0394 -		ISM02.4	PPB	9/1/2017 17:38	SMPL

INSTRUMENT ID : CV1

I5040-09	MA41E6	-264	-0.0587 -	ISM02.4	PPB	9/1/2017 17:40	SMPL
I5040-10	MA41E7	177	0.0479 -	ISM02.4	PPB	9/1/2017 17:43	SMPL
I5040-11	MA41E8	813	0.2017 -	ISM02.4	PPB	9/1/2017 17:45	SMPL
I5040-12	MA41E9	3001	0.7307 -	ISM02.4	PPB	9/1/2017 17:47	SMPL
I5040-13	MA41G0	-263	-0.0585 -	ISM02.4	PPB	9/1/2017 17:50	SMPL
I5040-14	MA41G1	1400	0.3436 -	ISM02.4	PPB	9/1/2017 17:52	SMPL
CCV023	CCV023	20318	4.9174 -	ISM02.4	PPB	9/1/2017 17:54	SMPL
CCB023	CCB023	-178	-0.0379 -	ISM02.4	PPB	9/1/2017 17:56	SMPL
I5040-15	MA41G2	-100	-0.019 -	ISM02.4	PPB	9/1/2017 17:59	SMPL
I5040-16	MA41G3	473	0.1195 -	ISM02.4	PPB	9/1/2017 18:01	SMPL
I5040-17	MA41G4	69	0.0218 -	ISM02.4	PPB	9/1/2017 18:03	SMPL
I5040-18	MA41G5	833075	HIGH -	ISM02.4	PPB	9/1/2017 18:05	
I5040-19	MA41G6	802334	HIGH -	ISM02.4	PPB	9/1/2017 18:12	
PB102011BL	PBW010	-332	-0.0751 -	ISM02.4	PPB	9/1/2017 18:25	SMPL
PB102011TB	LEB010	-190	-0.0408 -	ISM02.4	PPB	9/1/2017 18:28	SMPL
I4960-02	5-POST1	913	0.2259 -	ISM02.4	PPB	9/1/2017 18:30	SMPL
CCV024	CCV024	20505	4.9626 -	ISM02.4	PPB	9/1/2017 18:32	SMPL
CCB024	CCB024	-227	-0.0497 -	ISM02.4	PPB	9/1/2017 18:35	SMPL
I5040-18DLX10	MA41G5	155098	37.5031 -	ISM02.4	PPB	9/1/2017 18:37	SMPL
I5040-19DLX10	MA41G6	149708	36.2 -	ISM02.4	PPB	9/1/2017 18:44	SMPL
CCV025	CCV025	19520	4.7245 -	ISM02.4	PPB	9/1/2017 18:46	SMPL
CCB025	CCB025	-428	-0.0983 -	ISM02.4	PPB	9/1/2017 18:49	SMPL
I5040-18DL2X50	MA41G5	31556	7.6344 -	ISM02.4	PPB	9/1/2017 18:51	SMPL
I5040-19DL2X50	MA41G6	29465	7.1289 -	ISM02.4	PPB	9/1/2017 18:58	SMPL
CCV026	CCV026	20177	4.8833 -	ISM02.4	PPB	9/1/2017 19:00	SMPL
CCB026	CCB026	-161	-0.0338 -	ISM02.4	PPB	9/1/2017 19:03	SMPL