


7470A

LB89857
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36306



A= 0.0000e+000

B= 2.7368e-004

C= 1.0625e-002

Rho= 0.9999426

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	-135	0.000	-135				
0.2	0.200	0.192	-0.008	661	0.0 %	661				
2.5	2.500	2.492	-0.008	9065	0.0 %	9065				
5.0	5.000	5.071	0.071	18491	0.0 %	18491				
7.5	7.500	7.525	0.025	27458	0.0 %	27458				
10.0	10.000	9.947	-0.053	36306	0.0 %	36306				

LB89857 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Sep 2017 12:24:12				
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-135	-	0.0007470A	PPB	05 Sep 2017 12:29:01	S	Std
0.2 - 1	50.2	661	-	0.20007470A	PPB	05 Sep 2017 12:31:10	S	Std
2.5 - 1	52.5	9065	-	2.50007470A	PPB	05 Sep 2017 12:33:18	S	Std
5.0 - 1	55	18491	-	5.00007470A	PPB	05 Sep 2017 12:35:26	S	Std
7.5 - 1	57.5	27458	-	7.50007470A	PPB	05 Sep 2017 12:37:34	S	Std
10.0 - 1	510	36306	-	10.00007470A	PPB	05 Sep 2017 12:39:42	S	Std
ICV40 - 1	ICV40	14365	3.9420	-7470A	PPB	05 Sep 2017 12:42:16	U	SMPL
ICB40 - 1	ICB40	-311	-0.0745	-7470A	PPB	05 Sep 2017 12:44:22	U	SMPL
CCV29 - 1	CCV29	18504	5.0748	-7470A	PPB	05 Sep 2017 12:48:54	U	SMPL
CCB29 - 1	CCB29	-298	-0.0709	-7470A	PPB	05 Sep 2017 12:51:00	U	SMPL
CRA - 1	CRA	624	0.1814	-7470A	PPB	05 Sep 2017 12:53:08	U	SMPL
HighStd - 1	HighStd	36953	10.1239	-7470A	PPB	05 Sep 2017 12:55:14	U	SMPL
ChkStd - 1	ChkStd	25889	7.0959	-7470A	PPB	05 Sep 2017 12:57:21	U	SMPL
PB102029BL - 1	PBW	-368	-0.0901	-7470A	PPB	05 Sep 2017 12:59:30	U	SMPL
PB102029BS - 1	LCSW	14457	3.9672	-7470A	PPB	05 Sep 2017 13:01:39	U	SMPL
PB102029TB - 1	PB101979TB	-289	-0.0685	-7470A	PPB	05 Sep 2017 13:03:47	U	SMPL
I5036-03 - 1	T2-T3-COMP	-213	-0.0477	-7470A	PPB	05 Sep 2017 13:05:55	U	SMPL
I5049-03 - 1	TR-04-083017	-134	-0.0260	-7470A	PPB	05 Sep 2017 13:08:04	U	SMPL
I5050-04 - 1	RT4651-1540	-182	-0.0392	-7470A	PPB	05 Sep 2017 13:10:12	U	SMPL
I5055-06 - 1	RT3992-20-41-RT3451-1550	-55	-0.0044	-7470A	PPB	05 Sep 2017 13:12:18	U	SMPL
CCV30 - 1	CCV30	18658	5.1169	-7470A	PPB	05 Sep 2017 13:14:25	U	SMPL
CCB30 - 1	CCB30	-284	-0.0671	-7470A	PPB	05 Sep 2017 13:16:31	U	SMPL
I5068-03 - 1	HR-06-083117	-58	-0.0052	-7470A	PPB	05 Sep 2017 13:18:38	U	SMPL
I5069-03 - 1	I115	-217	-0.0488	-7470A	PPB	05 Sep 2017 13:20:44	U	SMPL
I5073-02 - 1	TP-1-2-COMP	-76	-0.0102	-7470A	PPB	05 Sep 2017 13:22:51	U	SMPL
I5073-09 - 1	TP-3-4-COMP	-167	-0.0351	-7470A	PPB	05 Sep 2017 13:24:58	U	SMPL
I5074-02 - 1	TP-5-6-COMP	-107	-0.0187	-7470A	PPB	05 Sep 2017 13:27:06	U	SMPL
I5074-09 - 1	TP-7-8-COMP	736	0.2121	-7470A	PPB	05 Sep 2017 13:29:13	U	SMPL
I5074-16 - 1	TP-9-10-COMP	-199	-0.0438	-7470A	PPB	05 Sep 2017 13:31:22	U	SMPL
I5074-16DUP - 1	TP-9-10-COMPDUP	-181	-0.0389	-7470A	PPB	05 Sep 2017 13:33:30	U	SMPL
I5074-16MS - 1	TP-9-10-COMPMS	13788	3.7841	-7470A	PPB	05 Sep 2017 13:35:39	U	SMPL
I5074-16MSD - 1	TP-9-10-COMPMSD	13805	3.7888	-7470A	PPB	05 Sep 2017 13:37:46	U	SMPL
CCV31 - 1	CCV31	18215	4.9957	-7470A	PPB	05 Sep 2017 13:39:53	U	SMPL
CCB31 - 1	CCB31	-318	-0.0764	-7470A	PPB	05 Sep 2017 13:42:00	U	SMPL
PB102030BL - 1	PBW	-189	-0.0411	-7470A	PPB	05 Sep 2017 13:44:07	U	SMPL
PB102030BS - 1	LCSW	13739	3.7707	-7470A	PPB	05 Sep 2017 13:46:14	U	SMPL
I5057-01 - 1	MWHR2-6	-481	-0.1210	-7470A	PPB	05 Sep 2017 13:48:21	U	SMPL
I5057-01DUP - 1	MWHR2-6DUP	-365	-0.0893	-7470A	PPB	05 Sep 2017 13:50:29	U	SMPL
I5057-02 - 1	I5057-01MS	13821	3.7931	-7470A	PPB	05 Sep 2017 13:52:36	U	SMPL
I5057-03 - 1	I5057-01MSD	14505	3.9803	-7470A	PPB	05 Sep 2017 13:54:43	U	SMPL
I5057-04 - 1	MWHR2-10	-457	-0.1144	-7470A	PPB	05 Sep 2017 13:56:51	U	SMPL
I5057-05 - 1	MWHR3-8	-238	-0.0545	-7470A	PPB	05 Sep 2017 13:58:59	U	SMPL
I5078-03 - 1	MWHR3-14	165	0.0558	-7470A	PPB	05 Sep 2017 14:01:08	U	SMPL
CCV32 - 1	CCV32	18483	5.0690	-7470A	PPB	05 Sep 2017 14:03:15	U	SMPL
CCB32 - 1	CCB32	-305	-0.0728	-7470A	PPB	05 Sep 2017 14:05:22	U	SMPL
I5074-16LX5 - 1		-195	-0.0427	-7470A	PPB	05 Sep 2017 14:07:30	U	SMPL
I5074-16A - 1		14121	3.8752	-7470A	PPB	05 Sep 2017 14:09:37	U	SMPL
I5057-01LX5 - 1		-311	-0.0745	-7470A	PPB	05 Sep 2017 14:11:44	U	SMPL
I5057-01A - 1		14381	3.9464	-7470A	PPB	05 Sep 2017 14:13:51	U	SMPL
CCV33 - 1	CCV33	18340	5.0299	-7470A	PPB	05 Sep 2017 14:15:57	U	SMPL
CCB33 - 1	CCB33	-301	-0.0718	-7470A	PPB	05 Sep 2017 14:18:05	U	SMPL