



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89860

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/05/2017	11:15
2	CAL2	1	Water			7.01	09/05/2017	11:16
3	CAL3	1	Water			10.04	09/05/2017	11:17
4	ICV	1	Water			7.00	09/05/2017	11:19
5	CCV1	1	Water			2.01	09/05/2017	11:22
6	I5052-13	1	Solid	20.02	20	8.20	09/05/2017	11:24
7	I5052-15	1	Solid	20.03	20	9.63	09/05/2017	11:25
8	I5052-17	1	Solid	20.04	20	9.31	09/05/2017	11:26
9	I5052-18	1	Solid	20.02	20	8.45	09/05/2017	11:27
10	I5053-09	1	Solid	20.03	20	10.47	09/05/2017	11:30
11	I5053-11	1	Solid	20.02	20	9.15	09/05/2017	11:31
12	I5053-12	1	Solid	20.04	20	8.62	09/05/2017	11:32
13	I5071-19	1	Solid	20.02	20	8.65	09/05/2017	11:34
14	I5071-20	1	Solid	20.03	20	9.46	09/05/2017	11:36
15	I5071-21	1	Solid	20.02	20	8.81	09/05/2017	11:37
16	CCV2	1	Water			12.02	09/05/2017	11:38
17	I5071-22	1	Solid	20.01	20	10.34	09/05/2017	11:39
18	I5071-23	1	Solid	20.02	20	9.63	09/05/2017	11:40
19	I5071-24	1	Solid	20.03	20	8.32	09/05/2017	11:41
20	I5071-25	1	Solid	20.02	20	7.86	09/05/2017	11:43

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I5071-26	1	Solid	20.03	20	9.76	09/05/2017	11:45
22	I5071-27	1	Solid	20.05	20	8.94	09/05/2017	11:46
23	I5071-27DUP	1	Solid	20.03	20	8.95	09/05/2017	11:47
24	CCV3	1	Water			12.02	09/05/2017	11:48

WORKLIST(Hardcopy Internal Chain)

1889860

WorkList Name : CORROSIONITY I5052

WorkList ID : 102554

Date : 9/5/2017 10:59:11 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5052-13	Corrosivity	Cool 4 deg C	TECT01	O51	G26A-0-3-COMP	09/01/2017	9045C
	Solid	I5052-15	Corrosivity	Cool 4 deg C	TECT01	O51	G30A-0-3-COMP	09/01/2017	9045C
	Solid	I5052-17	Corrosivity	Cool 4 deg C	TECT01	O51	G27A-0-1.5	08/30/2017	9045C
	Solid	I5052-18	Corrosivity	Cool 4 deg C	TECT01	O51	G27A-1.5-3	08/30/2017	9045C
	Solid	I5053-09	Corrosivity	Cool 4 deg C	TECT01	O51	G-46A-0-3-COMP	08/30/2017	9045C
	Solid	I5053-11	Corrosivity	Cool 4 deg C	TECT01	O51	G-62A-0-1.5	08/30/2017	9045C
	Solid	I5053-12	Corrosivity	Cool 4 deg C	TECT01	O51	G-62A-1.5-3	08/30/2017	9045C
	Solid	I5071-19	Corrosivity	Cool 4 deg C	TECT01	O53	G43A-1.5-3	08/31/2017	9045C
	Solid	I5071-20	Corrosivity	Cool 4 deg C	TECT01	O53	G61A-0-1.5	08/31/2017	9045C
	Solid	I5071-21	Corrosivity	Cool 4 deg C	TECT01	O53	G61A-1.5-3	08/31/2017	9045C
	Solid	I5071-22	Corrosivity	Cool 4 deg C	TECT01	O53	G58A-0-2.5	08/31/2017	9045C
	Solid	I5071-23	Corrosivity	Cool 4 deg C	TECT01	O53	G58A-2.5-5	08/31/2017	9045C
	Solid	I5071-24	Corrosivity	Cool 4 deg C	TECT01	O53	G55A-0-2.5	08/31/2017	9045C
	Solid	I5071-25	Corrosivity	Cool 4 deg C	TECT01	O53	G55A-2.5-4.5	08/31/2017	9045C
	Solid	I5071-26	Corrosivity	Cool 4 deg C	TECT01	O53	G57A-0-1.5	08/31/2017	9045C
	Solid	I5071-27	Corrosivity	Cool 4 deg C	TECT01	O53	G57A-1.5-3	08/31/2017	9045C

Date/Time 09-05-17 11:00 AM
 Received by: JB
 Relinquished by: JB

Date/Time 09-05-17 2:13 PM
 Received by: JB
 Relinquished by: JB