

7471A


LB89869

INSTRUMENT ID: CY1

Linear

 μ Abs.:

42953



A= 0.0000e+000

B= 2.3009e-004

C= 3.9578e-003

Rho= 0.9997531

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-68	0.000	-68				
0.2	0.200	0.202	0.002	862	0.0 %	862				
2.5	2.500	2.472	-0.028	10725	0.0 %	10725				
5.0	5.000	4.988	-0.012	21662	0.0 %	21662				
7.5	7.500	7.663	0.163	33286	0.0 %	33286	0			
10.0	10.000	9.887	-0.113	42953	0.0 %	42953				

LB89869 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 05 Sep 2017 15:10:01				
Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-68	-	0.00007471A	PPB	05 Sep 2017 15:12:30	S	Std
0.2 - 1	502	862	-	0.20007471A	PPB	05 Sep 2017 15:14:48	S	Std
2.5 - 1	525	10725	-	2.50007471A	PPB	05 Sep 2017 15:17:06	S	Std
5.0 - 1	55	21662	-	5.00007471A	PPB	05 Sep 2017 15:19:23	S	Std
7.5 - 1	575	33286	-	7.50007471A	PPB	05 Sep 2017 15:24:14	S	Std
10.0 - 1	510	42953	-	10.00007471A	PPB	05 Sep 2017 15:26:30	S	Std
ICV41 - 1	ICV41	16208	3.7332	-7471A	PPB	05 Sep 2017 15:29:24	U	SMPL
ICB41 - 1	ICB41	-289	-0.0625	-7471A	PPB	05 Sep 2017 15:31:40	U	SMPL
CCV34 - 1	CCV34	20881	4.8084	-7471A	PPB	05 Sep 2017 15:33:56	U	SMPL
CCB34 - 1	CCB34	-285	-0.0616	-7471A	PPB	05 Sep 2017 15:36:12	U	SMPL
CRA - 1	CRA	834	0.1959	-7471A	PPB	05 Sep 2017 15:38:28	U	SMPL
HighStd - 1	HighStd	42978	9.8927	-7471A	PPB	05 Sep 2017 15:40:44	U	SMPL
ChkStd - 1	ChkStd	29748	6.8486	-7471A	PPB	05 Sep 2017 15:43:00	U	SMPL
PB102042BL - 1	PBS	-113	-0.0220	-7471A	PPB	05 Sep 2017 15:45:19	U	SMPL
PB102042BS - 1	LCSS	15873	3.6561	-7471A	PPB	05 Sep 2017 15:47:37	U	SMPL
I5052-01 - 1	G26A-0-3-COMP	8754	2.0182	-7471A	PPB	05 Sep 2017 15:49:54	U	SMPL
I5052-06 - 1	G30A-0-3-COMP	5335	1.2315	-7471A	PPB	05 Sep 2017 15:52:11	U	SMPL
I5052-11 - 1	G27A-0-1.5	1307	0.3047	-7471A	PPB	05 Sep 2017 15:54:28	U	SMPL
I5052-12 - 1	G27A-1.5-3	9672	2.2294	-7471A	PPB	05 Sep 2017 15:56:46	U	SMPL
I5053-01 - 1	G-46A-0-3-COMP	7195	1.6594	-7471A	PPB	05 Sep 2017 15:59:02	U	SMPL
CCV35 - 1	CCV35	20690	4.7645	-7471A	PPB	05 Sep 2017 16:01:18	U	SMPL
CCB35 - 1	CCB35	-279	-0.0602	-7471A	PPB	05 Sep 2017 16:03:33	U	SMPL
I5053-05 - 1	G-62A-0-1.5	369	0.0889	-7471A	PPB	05 Sep 2017 16:05:49	U	SMPL
I5053-06 - 1	G-62A-1.5-3	4743	1.0953	-7471A	PPB	05 Sep 2017 16:08:05	U	SMPL
I5075-01 - 1	72-11948	3364	0.7780	-7471A	PPB	05 Sep 2017 16:10:21	U	SMPL
I5075-04 - 1	72-12020	5273	1.2172	-7471A	PPB	05 Sep 2017 16:12:38	U	SMPL
I5075-07 - 1	VNJ-221	22300	5.1349	-7471A	PPB	05 Sep 2017 16:14:55	U	SMPL
I5075-10 - 1	292	35806	8.2425	-7471A	PPB	05 Sep 2017 16:17:12	U	SMPL
I5076-01 - 1	OR-02-083117	1647	0.3829	-7471A	PPB	05 Sep 2017 16:19:29	U	SMPL
I5076-01DUP - 1	OR-02-083117DUP	1657	0.3852	-7471A	PPB	05 Sep 2017 16:21:48	U	SMPL
I5076-01MS - 1	OR-02-083117MS	17048	3.9265	-7471A	PPB	05 Sep 2017 16:24:06	U	SMPL
I5076-01MSD - 1	OR-02-083117MSD	16916	3.8961	-7471A	PPB	05 Sep 2017 16:26:22	U	SMPL
CCV36 - 1	CCV36	20458	4.7111	-7471A	PPB	05 Sep 2017 16:28:38	U	SMPL
CCB36 - 1	CCB36	-274	-0.0591	-7471A	PPB	05 Sep 2017 16:30:55	U	SMPL
I5083-01 - 1	CLEAN-SAND-8	462	0.1103	-7471A	PPB	05 Sep 2017 16:33:12	U	SMPL
I5083-02 - 1	CLEAN-SAND-9	710	0.1673	-7471A	PPB	05 Sep 2017 16:35:28	U	SMPL
I5084-01 - 1	TP9-WC1	11517	2.6539	-7471A	PPB	05 Sep 2017 16:37:44	U	SMPL
I5084-02 - 1	TP10-WC1	27951	6.4352	-7471A	PPB	05 Sep 2017 16:40:01	U	SMPL
I5092-01 - 1	BU-02-090117	1101	0.2573	-7471A	PPB	05 Sep 2017 16:42:17	U	SMPL
I5093-01 - 1	AU-06-090117-A	1086	0.2538	-7471A	PPB	05 Sep 2017 16:44:36	U	SMPL
I5093-04 - 1	AU-06-090117-B	1306	0.3045	-7471A	PPB	05 Sep 2017 16:46:53	U	SMPL
I5093-07 - 1	AU-06-090117-C	1139	0.2660	-7471A	PPB	05 Sep 2017 16:49:11	U	SMPL
PB102043BL - 1	PBS	237	0.0585	-7471A	PPB	05 Sep 2017 16:51:29	U	SMPL
PB102043BS - 1	LCSS	15315	3.5278	-7471A	PPB	05 Sep 2017 16:53:44	U	SMPL
CCV37 - 1	CCV37	20898	4.8123	-7471A	PPB	05 Sep 2017 16:56:01	U	SMPL
CCB37 - 1	CCB37	-264	-0.0568	-7471A	PPB	05 Sep 2017 16:58:17	U	SMPL
I5071-01 - 1	G43A-1.5-3	8977	2.0695	-7471A	PPB	05 Sep 2017 17:00:34	U	SMPL
I5071-03 - 1	G61A-0-1.5	2834	0.6560	-7471A	PPB	05 Sep 2017 17:02:50	U	SMPL
I5071-04 - 1	G61A-1.5-3	17631	4.0606	-7471A	PPB	05 Sep 2017 17:05:06	U	SMPL
I5071-07 - 1	G58A-0-2.5	1866	0.4333	-7471A	PPB	05 Sep 2017 17:07:23	U	SMPL
I5071-08 - 1	G58A-2.5-5	4824	1.1139	-7471A	PPB	05 Sep 2017 17:09:40	U	SMPL
I5071-13 - 1	G55A-0-2.5	19748	4.5477	-7471A	PPB	05 Sep 2017 17:12:02	U	SMPL
I5071-13DUP - 1	G55A-0-2.5DUP	19562	4.5049	-7471A	PPB	05 Sep 2017 17:14:18	U	SMPL
I5071-13MS - 1	G55A-0-2.5MS	38457	8.8525	-7471A	PPB	05 Sep 2017 17:16:34	U	SMPL
I5071-13MSD - 1	G55A-0-2.5MSD	39423	9.0747	-7471A	PPB	05 Sep 2017 17:18:50	U	SMPL
I5071-14 - 1	G55A-2.5-4.5	10669	2.4588	-7471A	PPB	05 Sep 2017 17:21:06	U	SMPL
CCV38 - 1	CCV38	21574	4.9679	-7471A	PPB	05 Sep 2017 17:23:22	U	SMPL
CCB38 - 1	CCB38	-310	-0.0674	-7471A	PPB	05 Sep 2017 17:25:38	U	SMPL
I5071-17 - 1	G57A-0-1.5	2959	0.6848	-7471A	PPB	05 Sep 2017 17:27:54	U	SMPL
I5071-18 - 1	G57A-1.5-3	8295	1.9125	-7471A	PPB	05 Sep 2017 17:30:10	U	SMPL
I5095-01 - 1	STORM-SEWER-SED	106225	24.4451	-7471A	PPB	05 Sep 2017 17:32:26	U	SMPL
I5076-01LX5 - 1		535	0.1271	-7471A	PPB	05 Sep 2017 17:34:42	U	SMPL
I5076-01A - 1		18713	4.3096	-7471A	PPB	05 Sep 2017 17:36:58	U	SMPL
I5071-13LX5 - 1		4576	1.0568	-7471A	PPB	05 Sep 2017 17:39:14	U	SMPL
I5071-13A - 1		34112	7.8527	-7471A	PPB	05 Sep 2017 17:41:30	U	SMPL
I5095-01DLX5 - 1		21383	4.9239	-7471A	PPB	05 Sep 2017 17:43:46	U	SMPL
CCV39 - 1	CCV39	21638	4.9826	-7471A	PPB	05 Sep 2017 17:46:02	U	SMPL
CCB39 - 1	CCB39	-275	-0.0593	-7471A	PPB	05 Sep 2017 17:48:18	U	SMPL