



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89870

Temp.: 19.8 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/05/2017	15:10
2	CAL2	1	Water			7.01	09/05/2017	15:11
3	CAL3	1	Water			10.03	09/05/2017	15:13
4	ICV	1	Water			6.99	09/05/2017	15:15
5	CCV1	1	Water			2.01	09/05/2017	15:16
6	I5103-04	1	Water			6.32	09/05/2017	15:17
7	I5103-04DUP	1	Water			6.33	09/05/2017	15:18
8	CCV2	1	Water			12.01	09/05/2017	15:20

LB89870

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5104

WorkList ID : 102571

Date : 9/5/2017 2:59:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5103-04 ✓	pH	Cool 4 deg C	MAJO01	O11	DRAIN-WATER-TANK-1	09/05/2017	SM 4500-PH B

Date/Time 09-05-17 2:00 PM
Received by: JO
Relinquished by: _____

Date/Time 09-05-17 4:00 PM
Received by: CJ
Relinquished by: JO