

ISM02.4


LB89881

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43954



A= 0.0000e+000

B= 2.2751e-004

C= 3.2686e-002

Rho= 0.9999666

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	-0.021	-0.021	-235	0.000	-235					6
0.2	0.200	0.213	0.013	792	0.0 %	792					1
2.5	2.500	2.534	0.034	10994	0.0 %	10994					0
5.0	5.000	4.992	-0.008	21797	0.0 %	21797					0
7.5	7.500	7.449	-0.051	32599	0.0 %	32599					0
10.0	10.000	10.033	0.033	43954	0.0 %	43954					0

ns
9/5/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		-235	-		0 ISM02.4	PPB	9/5/2017 18:29	Std
0.2 S0.2		792	-		0.2 ISM02.4	PPB	9/5/2017 18:32	Std
2.5 S2.5		10994	-		2.5 ISM02.4	PPB	9/5/2017 18:34	Std
5 S5		21797	-		5 ISM02.4	PPB	9/5/2017 18:36	Std
7.5 S7.5		32599	-		7.5 ISM02.4	PPB	9/5/2017 18:39	Std
10 S10		43954	-		10 ISM02.4	PPB	9/5/2017 18:41	Std
ICV	ICV	17045	3.9106 -		ISM02.4	PPB	9/5/2017 18:44	SMPL
ICB	ICB	-340	-0.0447 -		ISM02.4	PPB	9/5/2017 18:46	SMPL
CCV040	CCV040	21820	4.997 -		ISM02.4	PPB	9/5/2017 18:48	SMPL
CCB040	CCB040	-212	-0.0155 -		ISM02.4	PPB	9/5/2017 18:50	SMPL
PB102068BL	PBS004	-18	0.0286 -		ISM02.4	PPB	9/5/2017 18:53	SMPL
I5099-01	MBDZS5	2450	0.5901 -		ISM02.4	PPB	9/5/2017 18:55	SMPL
I5099-02	MBDZS5D	2864	0.6843 -		ISM02.4	PPB	9/5/2017 18:57	SMPL
I5099-03	MBDZS5S	13896	3.1942 -		ISM02.4	PPB	9/5/2017 18:59	SMPL
I5099-04	MBDZS6	2753	0.659 -		ISM02.4	PPB	9/5/2017 19:02	SMPL
I5099-05	MBDZS7	2601	0.6244 -		ISM02.4	PPB	9/5/2017 19:04	SMPL
I5099-06	MBDZT4	1249	0.3168 -		ISM02.4	PPB	9/5/2017 19:06	SMPL
I5099-07	MBDZT5	1013	0.2632 -		ISM02.4	PPB	9/5/2017 19:09	SMPL
I5099-08	MBDZT6	846	0.2252 -		ISM02.4	PPB	9/5/2017 19:11	SMPL
CCV041	CCV041	22017	5.0418 -		ISM02.4	PPB	9/5/2017 19:13	SMPL
CCB041	CCB041	-341	-0.0449 -		ISM02.4	PPB	9/5/2017 19:15	SMPL