



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89883

Temp.: 19.8 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/05/2017	17:50
2	CAL2	1	Water			7.01	09/05/2017	17:51
3	CAL3	1	Water			10.03	09/05/2017	17:52
4	ICV	1	Water			7.00	09/06/2017	17:53
5	CCV1	1	Water			2.01	09/05/2017	17:54
6	I5109-02	1	Water			7.15	09/05/2017	18:00
7	I5109-02DUP	1	Water			7.16	09/05/2017	18:01
8	CCV2	1	Water			12.01	09/05/2017	18:02

LB 89883

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I5109

WorkList ID : 102595

Date : 9/5/2017 5:37:10 pm

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2017	Water	I5109-02-13	Corrosivity	Cool 4 deg C	PSEG01		DRUM-1-2-COMP	09/05/2017	9040C

Date/Time 09-05-17 5:37 PM
Received by: JP
Relinquished by: JP

Date/Time 09-05-17 5:00 PM
Received by: JP
Relinquished by: JP