



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89884

Temp.: 19.8 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/05/2017	17:50
2	CAL2	1	Water			7.01	09/05/2017	17:51
3	CAL3	1	Water			10.03	09/05/2017	17:52
4	ICV	1	Water			7.00	09/05/2017	17:53
5	CCV1	1	Water			2.01	09/05/2017	17:54
6	I5109-01	1	Water			7.15	09/05/2017	18:00
7	I5109-01DUP	1	Water			7.16	09/05/2017	18:01
8	CCV2	1	Water			12.01	09/05/2017	18:02

1389 884

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i5109

WorkList ID : 102596

Date : 9/5/2017 5:38:10 pm

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/05/2017	Water	I5109-01-13	pH	Cool 4 deg C	PSEG01	O52	DRUM-1-2-COMP	09/05/2017	9045C

Date/Time 09-05-17 5:37 PM
Received by: JP
Relinquished by: CG

Date/Time 09-05-17 7:00 PM
Received by: CD
Relinquished by: Jo