



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89885

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/05/2017	17:50
2	CAL2	1	Water			7.01	09/05/2017	17:52
3	CAL3	1	Water			10.03	09/05/2017	17:53
4	ICV	1	Water			7.00	09/05/2017	17:54
5	CCV1	1	Water			2.02	09/05/2017	18:10
6	I5105-03	1	Solid	20.02	20	8.10	09/05/2017	18:11
7	I5110-03	1	Solid	20.03	20	8.20	09/05/2017	18:13
8	I5110-06	1	Solid	20.02	20	8.63	09/05/2017	18:15
9	I5110-06DUP	1	Solid	20.03	20	8.64	09/05/2017	18:17
10	CCV2	1	Water			12.02	09/05/2017	18:20

1889885

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5105

WorkList ID : 102597

Date : 9/5/2017 5:42:25 pm

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/08/2017	Solid	I5105-03	Corrosivity	Cool 4 deg C	PSEG05	O11	OK-02-090517	09/05/2017	9045C
09/08/2017	Solid	I5110-03	Corrosivity	Cool 4 deg C	PSEG01	O52	SP-1	09/05/2017	9045C
09/08/2017	Solid	I5110-06	Corrosivity	Cool 4 deg C	PSEG01	O52	SP-2	09/05/2017	9045C

Date/Time 09-05-17 5:40 PM
 Received by: SP
 Relinquished by: CP

Date/Time 09-05-17 7:10 PM
 Received by: CP
 Relinquished by: JP