


7470A

LB89888
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35124



A= 0.0000e+000

B= 2.8487e-004

C= 1.5454e-002

Rho= 0.9999896

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	-34	0.000	-34				
0.2	0.200	0.194	-0.006	628	0.0 %	628				
2.5	2.500	2.510	0.010	8756	0.0 %	8756				
5.0	5.000	5.002	0.002	17503	0.0 %	17503				
7.5	7.500	7.467	-0.033	26158	0.0 %	26158				
10.0	10.000	10.021	0.021	35124	0.0 %	35124				

LB89888 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 06 Sep 2017 11:54:10					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-34	-	0.0000	7470A	PPB	06 Sep 2017 12:01:18	S	Std
0.2 - 1	50.2	628	-	0.2000	7470A	PPB	06 Sep 2017 12:03:27	S	Std
2.5 - 1	50.5	8756	-	2.5000	7470A	PPB	06 Sep 2017 12:05:35	S	Std
5.0 - 1	55	17503	-	5.0000	7470A	PPB	06 Sep 2017 12:07:44	S	Std
7.5 - 1	57.5	26158	-	7.5000	7470A	PPB	06 Sep 2017 12:09:51	S	Std
10.0 - 1	510	35124	-	10.0000	7470A	PPB	06 Sep 2017 12:11:59	S	Std
ICV42 - 1	ICV42	13552	3.8761	-	7470A	PPB	06 Sep 2017 12:14:27	U	SMPL
ICB42 - 1	ICB42	-365	-0.0885	-	7470A	PPB	06 Sep 2017 12:16:35	U	SMPL
CCV42 - 1	CCV42	17409	4.9748	-	7470A	PPB	06 Sep 2017 12:18:44	U	SMPL
CCB42 - 1	CCB42	-251	-0.0560	-	7470A	PPB	06 Sep 2017 12:20:50	U	SMPL
CRA - 1	CRA	574	0.1790	-	7470A	PPB	06 Sep 2017 12:22:57	U	SMPL
HighStd - 1	HighStd	35824	10.2207	-	7470A	PPB	06 Sep 2017 12:25:04	U	SMPL
ChkStd - 1	ChkStd	24917	7.1136	-	7470A	PPB	06 Sep 2017 12:27:11	U	SMPL
PB102081BL - 1	PBW	-234	-0.0512	-	7470A	PPB	06 Sep 2017 12:29:19	U	SMPL
PB102081BS - 1	LCSW	13396	3.8316	-	7470A	PPB	06 Sep 2017 12:31:29	U	SMPL
PB102081TB - 1	PB102025TB	-156	-0.0290	-	7470A	PPB	06 Sep 2017 12:33:36	U	SMPL
I5075-03 - 1	72-11948	-75	-0.0059	-	7470A	PPB	06 Sep 2017 12:35:45	U	SMPL
I5075-06 - 1	72-12020	-201	-0.0418	-	7470A	PPB	06 Sep 2017 12:37:53	U	SMPL
I5075-09 - 1	VNJ-221	-161	-0.0304	-	7470A	PPB	06 Sep 2017 12:40:01	U	SMPL
I5075-12 - 1	292	-150	-0.0273	-	7470A	PPB	06 Sep 2017 12:42:08	U	SMPL
CCV43 - 1	CCV43	17506	5.0024	-	7470A	PPB	06 Sep 2017 12:44:15	U	SMPL
CCB43 - 1	CCB43	-271	-0.0617	-	7470A	PPB	06 Sep 2017 12:46:22	U	SMPL
I5076-03 - 1	OR-02-083117	-86	-0.0090	-	7470A	PPB	06 Sep 2017 12:48:28	U	SMPL
I5084-03 - 1	TP9-WC1	-27	0.0078	-	7470A	PPB	06 Sep 2017 12:50:35	U	SMPL
I5084-04 - 1	TP10-WC1	1260	0.3744	-	7470A	PPB	06 Sep 2017 12:52:42	U	SMPL
I5086-04 - 1	SP-1	-67	-0.0036	-	7470A	PPB	06 Sep 2017 12:54:49	U	SMPL
I5086-04DUP - 1	SP-1DUP	-101	-0.0133	-	7470A	PPB	06 Sep 2017 12:56:57	U	SMPL
I5086-04MS - 1	SP-1MS	13606	3.8914	-	7470A	PPB	06 Sep 2017 12:59:04	U	SMPL
I5086-04MSD - 1	SP-1MSD	13253	3.7909	-	7470A	PPB	06 Sep 2017 13:01:13	U	SMPL
PB102082BL - 1	PBW	-158	-0.0296	-	7470A	PPB	06 Sep 2017 13:03:21	U	SMPL
PB102082BS - 1	LCSW	13137	3.7578	-	7470A	PPB	06 Sep 2017 13:05:29	U	SMPL
PB102082TB - 1	PB102034TB	-295	-0.0686	-	7470A	PPB	06 Sep 2017 13:07:36	U	SMPL
CCV44 - 1	CCV44	17520	5.0064	-	7470A	PPB	06 Sep 2017 13:09:43	U	SMPL
CCB44 - 1	CCB44	-348	-0.0837	-	7470A	PPB	06 Sep 2017 13:11:49	U	SMPL
I5092-03 - 1	BU-02-090117	-171	-0.0333	-	7470A	PPB	06 Sep 2017 13:13:56	U	SMPL
I5093-03 - 1	AU-06-090117-A	-75	-0.0059	-	7470A	PPB	06 Sep 2017 13:16:03	U	SMPL
I5093-06 - 1	AU-06-090117-B	-60	-0.0016	-	7470A	PPB	06 Sep 2017 13:18:10	U	SMPL
I5093-09 - 1	AU-06-090117-C	-52	0.0006	-	7470A	PPB	06 Sep 2017 13:20:17	U	SMPL
I5095-03 - 1	STORM-SEWER-SED	-24	0.0086	-	7470A	PPB	06 Sep 2017 13:22:25	U	SMPL
I5095-03DUP - 1	STORM-SEWER-SEDDUP	22	0.0217	-	7470A	PPB	06 Sep 2017 13:24:32	U	SMPL
I5095-03MS - 1	STORM-SEWER-SEDMS	13307	3.8063	-	7470A	PPB	06 Sep 2017 13:26:40	U	SMPL
I5095-03MSD - 1	STORM-SEWER-SEDMSD	13627	3.8974	-	7470A	PPB	06 Sep 2017 13:28:48	U	SMPL
CCV45 - 1	CCV45	17512	5.0041	-	7470A	PPB	06 Sep 2017 13:30:57	U	SMPL
CCB45 - 1	CCB45	-322	-0.0763	-	7470A	PPB	06 Sep 2017 13:33:05	U	SMPL
I5086-04LX5 - 1		-211	-0.0447	-	7470A	PPB	06 Sep 2017 13:35:12	U	SMPL
I5086-04A - 1		14012	4.0071	-	7470A	PPB	06 Sep 2017 13:37:19	U	SMPL
I5095-03LX5 - 1		-239	-0.0526	-	7470A	PPB	06 Sep 2017 13:39:26	U	SMPL
I5095-03A - 1		14594	4.1729	-	7470A	PPB	06 Sep 2017 13:41:33	U	SMPL
CCV46 - 1	CCV46	17926	5.1221	-	7470A	PPB	06 Sep 2017 13:43:41	U	SMPL
CCB46 - 1	CCB46	-211	-0.0447	-	7470A	PPB	06 Sep 2017 13:45:48	U	SMPL