



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89893

Temp.: 22.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/06/2017	14:10
2	CAL2	1	Water			7.00	09/06/2017	14:11
3	CAL3	1	Water			10.03	09/06/2017	14:12
4	ICV	1	Water			6.99	09/06/2017	14:15
5	CCV1	1	Water			2.01	09/06/2017	14:16
6	I5117-04	1	Water			5.47	09/06/2017	14:18
7	I5117-04DUP	1	Water			5.48	09/06/2017	14:20
8	CCV2	1	Water			12.01	09/06/2017	14:21

2889893

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5117 WorkList ID : 102604 Date : 9/6/2017 1:54:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5117-04 ✓	pH	Cool 4 deg C	MAJO01	D12	DRAIN-WATER-TANK-1	09/06/2017	SM 4500-PH B

Date/Time 09.06.17 1:54 PM
Received by: JP
Relinquished by: CR

Date/Time 09.06.17 3:20 PM
Received by: CR
Relinquished by: JO