



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB89895

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 5N sulfuric acid, Hexavalent Chromium ICV Std., etc.

Intercept: 0.0005 Slope: 0.779 Regression: 0.999965

Table with 12 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml/gm), Final Vol (ml), pH HN03, pH H2SO4, Absorb. at 540 (Backgrnd, Color), Absorbance Difference, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Indiana

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: apatel

Run Number: LB89895

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb. at 540		Absorbance Difference	Result	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.11	0.000	0.393	0.393	0.504	09/06/2017	13:33
2	ICB		1	100	100		2.45	0.000	0.000	0.000	-0.001	09/06/2017	13:34
3	CCV1	0.5	1	100	100		1.61	0.000	0.394	0.394	0.505	09/06/2017	13:34
4	CCB1		1	100	100		1.88	0.000	0.000	0.000	-0.001	09/06/2017	13:35
5	PB102105BL		1	2.53	100	7.20	2.30	0.000	0.000	0.000	-0.001	09/06/2017	13:35
6	PB102105BS	20	1	2.51	100	7.39	2.14	0.000	0.391	0.391	0.501	09/06/2017	13:36
7	I5094-01		1	2.50	100	7.44	1.99	0.023	0.023	0.000	-0.001	09/06/2017	13:36
8	I5094-01DUP		1	2.55	100	7.78	2.23	0.022	0.022	0.000	-0.001	09/06/2017	13:37
9	I5094-02Pre	40	2	2.52	100	7.91	2.35	0.025	0.105	0.080	0.102	09/06/2017	13:37
10	I5094-022Ins	1284	40	2.54	100	7.85	1.77	0.024	0.629	0.605	0.776	09/06/2017	13:38
11	I5094-023Post	40	2	2.55	100	7.47	1.80	0.027	0.141	0.114	0.146	09/06/2017	13:38
12	I5094-03	40	2	2.52	100	7.91	2.35	0.025	0.105	0.080	0.102	09/06/2017	13:39
13	I5094-04		1	2.50	100	7.21	2.44	0.040	0.040	0.000	-0.001	09/06/2017	13:39
14	I5094-05		1	2.54	100	7.32	2.39	0.060	0.060	0.000	-0.001	09/06/2017	13:40
15	CCV2	0.5	1	100	100		1.79	0.000	0.394	0.394	0.505	09/06/2017	13:40
16	CCB2		1	100	100		1.58	0.000	0.000	0.000	-0.001	09/06/2017	13:41
17	I5094-09		1	2.51	100	7.66	2.24	0.035	0.035	0.000	-0.001	09/06/2017	13:41
18	I5094-10		1	2.52	100	7.04	2.41	0.030	0.030	0.000	-0.001	09/06/2017	13:42
19	I5105-01		1	2.51	100	7.11	1.60	0.020	0.020	0.000	-0.001	09/06/2017	13:42
20	I5110-01		1	2.50	100	7.23	1.87	0.028	0.028	0.000	-0.001	09/06/2017	13:43
21	I5110-04		1	2.55	100	7.46	2.20	0.018	0.018	0.000	-0.001	09/06/2017	13:43
22	I5113-01		1	2.50	100	7.25	2.10	0.015	0.015	0.000	-0.001	09/06/2017	13:44
23	I5113-03		1	2.54	100	7.80	2.47	0.037	0.037	0.000	-0.001	09/06/2017	13:44
24	I5113-05		1	2.53	100	7.88	1.77	0.041	0.041	0.000	-0.001	09/06/2017	13:45
25	I5113-07		1	2.55	100	7.90	1.93	0.059	0.059	0.000	-0.001	09/06/2017	13:45
26	CCV3	0.5	1	100	100		2.05	0.000	0.394	0.394	0.505	09/06/2017	13:46
27	CCB3		1	100	100		1.59	0.000	0.000	0.000	-0.001	09/06/2017	13:46