



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89902

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/06/2017	18:40
2	CAL2	1	Water			7.00	09/06/2017	18:41
3	CAL3	1	Water			10.03	09/06/2017	18:42
4	ICV	1	Water			7.01	09/06/2017	18:43
5	CCV1	1	Water			2.01	09/06/2017	18:44
6	I5120-03	1	Solid	20.02	20	8.33	09/06/2017	18:46
7	I5122-03	1	Solid	20.01	20	6.60	09/06/2017	18:47
8	I5126-03	1	Solid	20.03	20	8.62	09/06/2017	18:49
9	I5133-02	1	Solid	20.02	20	8.53	09/06/2017	18:50
10	I5133-02DUP	1	Solid	20.03	20	8.54	09/06/2017	18:51
11	CCV2	1	Water			12.02	09/06/2017	18:53

2389902

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5120

WorkList ID : 102610

Date : 9/6/2017 6:28:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5120-03	Corrosivity	Cool 4 deg C	WSP001	D22	B-16(0-10)COMP	09/06/2017	9045C
09/11/2017	Solid	I5122-03	Corrosivity	Cool 4 deg C	PSEG05	D52	SU-04-090617	09/06/2017	9045C
09/11/2017	Solid	I5126-03	Corrosivity	Cool 4 deg C	PSEG05	D52	TR-05-090617	09/06/2017	9045C
09/11/2017	Solid	I5133-02	Corrosivity	Cool 4 deg C	PSEG01	D52	VNJ-204	09/06/2017	9045C

Date/Time 090617-6309
Received by: JR
Relinquished by: JR

Date/Time 09-06-17 7:35PM
Received by: JR
Relinquished by: JR