


7470A

LB89905
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36310



A= 0.0000e+000

B= 2.7527e-004

C= -2.2755e-002

Rho= 0.9998768

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.036	-0.036	-48	0.000	-48				
0.2	0.200	0.192	-0.008	781	0.0 %	781				
2.5	2.500	2.493	-0.007	9138	0.0 %	9138				
5.0	5.000	5.126	0.126	18703	0.0 %	18703				
7.5	7.500	7.453	-0.047	27159	0.0 %	27159				
10.0	10.000	9.972	-0.028	36310	0.0 %	36310				

LB89905 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Sep 2017 09:50:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-48	-	0.0000	7470A	PPB	07 Sep 2017 09:57:37	S	Std
0.2 - 1	50.2	781	-	0.2000	7470A	PPB	07 Sep 2017 09:59:46	S	Std
2.5 - 1	52.5	9138	-	2.5000	7470A	PPB	07 Sep 2017 10:01:55	S	Std
5.0 - 1	55	18703	-	5.0000	7470A	PPB	07 Sep 2017 10:04:03	S	Std
7.5 - 1	57.5	27159	-	7.5000	7470A	PPB	07 Sep 2017 10:06:11	S	Std
10.0 - 1	510	36310	-	10.0000	7470A	PPB	07 Sep 2017 10:08:19	S	Std
ICV43 - 1	ICV43	13759	3.7647	-7470A	PPB	07 Sep 2017 10:11:02	U	SMPL	
ICB43 - 1	ICB43	-271	-0.0974	-7470A	PPB	07 Sep 2017 10:13:09	U	SMPL	
CCV47 - 1	CCV47	17931	4.9131	-7470A	PPB	07 Sep 2017 10:15:16	U	SMPL	
CCB47 - 1	CCB47	-200	-0.0778	-7470A	PPB	07 Sep 2017 10:17:23	U	SMPL	
CRA - 1	CRA	731	0.1785	-7470A	PPB	07 Sep 2017 10:19:30	U	SMPL	
HighStd - 1	HighStd	35879	9.8536	-7470A	PPB	07 Sep 2017 10:21:36	U	SMPL	
ChkStd - 1	ChkStd	24568	6.7400	-7470A	PPB	07 Sep 2017 10:23:43	U	SMPL	
PB102103BL - 1	PBW	-228	-0.0855	-7470A	PPB	07 Sep 2017 10:25:52	U	SMPL	
PB102103BS - 1	LCSW	13302	3.6389	-7470A	PPB	07 Sep 2017 10:28:01	U	SMPL	
PB102103TB - 1	PB102083TB	-248	-0.0910	-7470A	PPB	07 Sep 2017 10:30:09	U	SMPL	
I5052-13 - 1	G26A-0-3-COMP	-128	-0.0580	-7470A	PPB	07 Sep 2017 10:32:17	U	SMPL	
I5052-15 - 1	G30A-0-3-COMP	-82	-0.0453	-7470A	PPB	07 Sep 2017 10:34:25	U	SMPL	
I5052-17 - 1	G27A-0-1.5	-89	-0.0473	-7470A	PPB	07 Sep 2017 10:36:34	U	SMPL	
I5052-18 - 1	G27A-1.5-3	-98	-0.0497	-7470A	PPB	07 Sep 2017 10:38:41	U	SMPL	
CCV48 - 1	CCV48	18144	4.9717	-7470A	PPB	07 Sep 2017 10:40:48	U	SMPL	
CCB48 - 1	CCB48	-150	-0.0640	-7470A	PPB	07 Sep 2017 10:42:54	U	SMPL	
I5053-09 - 1	G-46A-0-3-COMP	-165	-0.0682	-7470A	PPB	07 Sep 2017 10:45:02	U	SMPL	
I5053-11 - 1	G-62A-0-1.5	-149	-0.0638	-7470A	PPB	07 Sep 2017 10:47:09	U	SMPL	
I5053-12 - 1	G-62A-1.5-3	-97	-0.0495	-7470A	PPB	07 Sep 2017 10:49:15	U	SMPL	
I5071-19 - 1	G43A-1.5-3	-65	-0.0406	-7470A	PPB	07 Sep 2017 10:51:24	U	SMPL	
I5071-20 - 1	G61A-0-1.5	-70	-0.0420	-7470A	PPB	07 Sep 2017 10:53:31	U	SMPL	
I5071-21 - 1	G61A-1.5-3	-58	-0.0387	-7470A	PPB	07 Sep 2017 10:55:39	U	SMPL	
I5071-22 - 1	G58A-0-2.5	-75	-0.0434	-7470A	PPB	07 Sep 2017 10:57:47	U	SMPL	
I5071-23 - 1	G58A-2.5-5	545	0.1273	-7470A	PPB	07 Sep 2017 10:59:55	U	SMPL	
I5071-24 - 1	G55A-0-2.5	-114	-0.0541	-7470A	PPB	07 Sep 2017 11:02:04	U	SMPL	
I5071-25 - 1	G55A-2.5-4.5	-82	-0.0453	-7470A	PPB	07 Sep 2017 11:04:11	U	SMPL	
CCV49 - 1	CCV49	17929	4.9125	-7470A	PPB	07 Sep 2017 11:06:18	U	SMPL	
CCB49 - 1	CCB49	-135	-0.0599	-7470A	PPB	07 Sep 2017 11:08:24	U	SMPL	
I5071-26 - 1	G57A-0-1.5	-81	-0.0451	-7470A	PPB	07 Sep 2017 11:10:31	U	SMPL	
I5071-27 - 1	G57A-1.5-3	-87	-0.0467	-7470A	PPB	07 Sep 2017 11:12:38	U	SMPL	
I5105-03 - 1	OK-02-090517	-121	-0.0561	-7470A	PPB	07 Sep 2017 11:14:45	U	SMPL	
I5109-02 - 1	DRUM-1-2-COMP	-102	-0.0508	-7470A	PPB	07 Sep 2017 11:16:52	U	SMPL	
I5110-03 - 1	SP-1	-90	-0.0475	-7470A	PPB	07 Sep 2017 11:18:59	U	SMPL	
I5110-06 - 1	SP-2	-61	-0.0395	-7470A	PPB	07 Sep 2017 11:21:07	U	SMPL	
I5110-06DUP - 1	SP-2DUP	-71	-0.0423	-7470A	PPB	07 Sep 2017 11:23:15	U	SMPL	
I5110-06MS - 1	SP-2MS	13156	3.5987	-7470A	PPB	07 Sep 2017 11:25:23	U	SMPL	
I5110-06MSD - 1	SP-2MSD	13250	3.6246	-7470A	PPB	07 Sep 2017 11:27:32	U	SMPL	
PB102104BL - 1	PBW	-181	-0.0726	-7470A	PPB	07 Sep 2017 11:29:40	U	SMPL	
CCV50 - 1	CCV50	18123	4.9659	-7470A	PPB	07 Sep 2017 11:31:47	U	SMPL	
CCB50 - 1	CCB50	-139	-0.0610	-7470A	PPB	07 Sep 2017 11:33:54	U	SMPL	
PB102104BS - 1	LCSW	13100	3.5833	-7470A	PPB	07 Sep 2017 11:36:02	U	SMPL	
I5109-01 - 1	DRUM-1-2-COMP	2188	0.5795	-7470A	PPB	07 Sep 2017 11:38:10	U	SMPL	
I5109-01MS - 1	DRUM-1-2-COMPMS	15587	4.2679	-7470A	PPB	07 Sep 2017 11:42:24	U	SMPL	
I5109-01MSD - 1	DRUM-1-2-COMPMSD	15393	4.2145	-7470A	PPB	07 Sep 2017 11:44:32	U	SMPL	
I5110-06LX5 - 1		-172	-0.0701	-7470A	PPB	07 Sep 2017 11:46:39	U	SMPL	
I5110-06A - 1		14545	3.9810	-7470A	PPB	07 Sep 2017 11:48:48	U	SMPL	
I5109-01LX5 - 1		313	0.0634	-7470A	PPB	07 Sep 2017 11:50:56	U	SMPL	
I5109-01A - 1		16248	4.4498	-7470A	PPB	07 Sep 2017 11:53:04	U	SMPL	
I5109-01DUP - 1	DRUM-1-2-COMP DUP	2032	0.5366	-7470A	PPB	07 Sep 2017 12:03:28	U	SMPL	
CCV51 - 1	CCV51	18022	4.9381	-7470A	PPB	07 Sep 2017 12:05:35	U	SMPL	
CCB51 - 1	CCB51	-159	-0.0665	-7470A	PPB	07 Sep 2017 12:07:42	U	SMPL	