



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89910

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	08/30/2017	14:00
2	CAL2	1	Water			7.01	08/30/2017	14:01
3	CAL3	1	Water			10.03	08/30/2017	14:02
4	ICV	1	Water			7.00	08/30/2017	14:03
5	CCV1	1	Water			2.02	08/30/2017	14:05
6	I4972-02	1	Solid	20.02	20	6.07	08/30/2017	14:06
7	I5026-03	1	Solid	20.03	20	10.76	08/30/2017	14:07
8	I5026-03DUP	1	Solid	20.05	20	10.77	08/30/2017	14:10
9	CCV2	1	Water			12.01	08/30/2017	14:11

WORKLIST(Hardcopy Internal Chain)

2889910

WorkList Name : CORROSIONITY I5026

WorkList ID : 102419

Date : 8/30/2017 1:42:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/29/2017	Solid	I4972-02	Corrosivity	Cool 4 deg C	ARCA06	O32	5-POST1	08/23/2017	9045C
	Solid	I5026-03	Corrosivity	Cool 4 deg C	WSP001	O52	B-17(0-10)COMP	08/29/2017	9045C

Date/Time 08/30/17 11:42 AM
 Received by: JS
 Relinquished by: JS

Date/Time 08/30/17 3:10 PM
 Received by: CD
 Relinquished by: JS