


245.1

LB89911
INSTRUMENT ID: CV1

Linear

 μ Abs.:

31037



A= 0.0000e+000

B= 3.2072e-004

C= 1.3909e-002

Rho= 0.9999452

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-41	0.000	-41				
0.2	0.200	0.224	0.024	655	0.0 %	655				
2.5	2.500	2.464	-0.036	7640	0.0 %	7640				
5.0	5.000	4.971	-0.029	15456	0.0 %	15456				
7.5	7.500	7.572	0.072	23567	0.0 %	23567				
10.0	10.000	9.968	-0.032	31037	0.0 %	31037				

LB89911 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Sep 2017 12:23:07

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000245.1	PPB	07 Sep 2017 12:27:25	S	Std
0.2 - 1	50.2	655	-	0.2000245.1	PPB	07 Sep 2017 12:29:34	S	Std
2.5 - 1	52.5	7640	-	2.5000245.1	PPB	07 Sep 2017 12:31:42	S	Std
5.0 - 1	55	15456	-	5.0000245.1	PPB	07 Sep 2017 12:33:50	S	Std
7.5 - 1	57.5	23567	-	7.5000245.1	PPB	07 Sep 2017 12:35:58	S	Std
10.0 - 1	510	31037	-	10.0000245.1	PPB	07 Sep 2017 12:39:22	S	Std
ICV44 - 1	ICV44	11839	3.8109	-245.1	PPB	07 Sep 2017 12:42:00	U	SMPL
ICB44 - 1	ICB44	-183	-0.0448	-245.1	PPB	07 Sep 2017 12:44:06	U	SMPL
CCV52 - 1	CCV52	15213	4.8930	-245.1	PPB	07 Sep 2017 12:46:13	U	SMPL
CCB52 - 1	CCB52	-170	-0.0406	-245.1	PPB	07 Sep 2017 12:48:19	U	SMPL
CRA - 1	CRA	614	0.2108	-245.1	PPB	07 Sep 2017 12:50:27	U	SMPL
HighStd - 1	HighStd	30748	9.8753	-245.1	PPB	07 Sep 2017 12:52:33	U	SMPL
ChkStd - 1	ChkStd	21119	6.7871	-245.1	PPB	07 Sep 2017 12:54:40	U	SMPL
PB102102BL - 1	PBW	-181	-0.0441	-245.1	PPB	07 Sep 2017 12:56:49	U	SMPL
PB102102BS - 1	LCSW	12900	4.1511	-245.1	PPB	07 Sep 2017 12:58:57	U	SMPL
I5088-01 - 1	083017-TIDO-F-D-M	-383	-0.1089	-245.1	PPB	07 Sep 2017 13:01:05	U	SMPL
I5088-01DUP - 1	083017-TIDO-F-D-MDUP	-349	-0.0980	-245.1	PPB	07 Sep 2017 13:03:12	U	SMPL
I5088-01MS - 1	083017-TIDO-F-D-MMS	14922	4.7996	-245.1	PPB	07 Sep 2017 13:05:20	U	SMPL
I5088-01MSD - 1	083017-TIDO-F-D-MMSD	14517	4.6697	-245.1	PPB	07 Sep 2017 13:07:29	U	SMPL
I5088-02 - 1	083017-TIDO-F-U-M	-409	-0.1173	-245.1	PPB	07 Sep 2017 13:09:35	U	SMPL
CCV53 - 1	CCV53	15386	4.9484	-245.1	PPB	07 Sep 2017 13:11:42	U	SMPL
CCB53 - 1	CCB53	-234	-0.0611	-245.1	PPB	07 Sep 2017 13:13:49	U	SMPL
I5089-01 - 1	083117-TIDO-E-D-M	-295	-0.0807	-245.1	PPB	07 Sep 2017 13:15:56	U	SMPL
I5089-02 - 1	083117-TIDO-E-U-M	-301	-0.0826	-245.1	PPB	07 Sep 2017 13:18:03	U	SMPL
I5090-01 - 1	090117-TIDO-E-D-M	-314	-0.0868	-245.1	PPB	07 Sep 2017 13:20:10	U	SMPL
I5090-02 - 1	090117-TIDO-E-U-M	-298	-0.0817	-245.1	PPB	07 Sep 2017 13:22:17	U	SMPL
I5088-01LX5 - 1		-199	-0.0499	-245.1	PPB	07 Sep 2017 13:24:25	U	SMPL
I5088-01A - 1		12237	3.9385	-245.1	PPB	07 Sep 2017 13:26:32	U	SMPL
CCV54 - 1	CCV54	15221	4.8955	-245.1	PPB	07 Sep 2017 13:28:40	U	SMPL
CCB54 - 1	CCB54	-150	-0.0342	-245.1	PPB	07 Sep 2017 13:30:48	U	SMPL