



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89926

Temp.: 20.1 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/07/2017	14:30
2	CAL2	1	Water			7.00	09/07/2017	14:31
3	CAL3	1	Water			10.02	09/07/2017	14:32
4	ICV	1	Water			7.00	09/07/2017	14:35
5	CCV1	1	Water			2.02	09/07/2017	14:36
6	I5152-04	1	Water			5.33	09/07/2017	14:37
7	I5152-04DUP	1	Water			5.34	09/07/2017	14:38
8	CCV2	1	Water			12.01	09/07/2017	14:39

2B 89926

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5152

WorkList ID : 102651

Date : 9/7/2017 2:22:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5152-04	pH	Cool 4 deg C	MAJO01	D11	DRAIN-WATER-TANK-1	09/07/2017	SM 4500-PH B

Date/Time 09.07.17 2:22:58 PM
 Received by: JA
 Relinquished by: JA

Date/Time 09.07.17 3:15 PM
 Received by: CP
 Relinquished by: JA