


7471A

LB89927
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42968



A= 0.0000e+000

B= 2.3366e-004

C= 2.1551e-002

Rho= 0.9999401

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	-29	0.000	-29				
0.2	0.200	0.212	0.012	817	0.0 %	817				
2.5	2.500	2.500	-0.000	10607	0.0 %	10607				
5.0	5.000	4.985	-0.015	21241	0.0 %	21241				
7.5	7.500	7.427	-0.073	31691	0.0 %	31691	0			
10.0	10.000	10.062	0.062	42968	0.0 %	42968				

LB89927 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Sep 2017 15:54:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7471A	PPB	07 Sep 2017 15:57:09	S	Std
0.2 - 1	50.2	817	-	0.2000	7471A	PPB	07 Sep 2017 15:59:27	S	Std
2.5 - 1	52.5	10607	-	2.5000	7471A	PPB	07 Sep 2017 16:01:45	S	Std
5.0 - 1	55	21241	-	5.0000	7471A	PPB	07 Sep 2017 16:04:02	S	Std
7.5 - 1	57.5	31691	-	7.5000	7471A	PPB	07 Sep 2017 16:08:54	S	Std
10.0 - 1	510	42968	-	10.0000	7471A	PPB	07 Sep 2017 16:11:10	S	Std
ICV45 - 1	ICV45	16679	3.9188	-	7471A	PPB	07 Sep 2017 16:13:49	U	SMPL
ICB45 - 1	ICB45	-292	-0.0467	-	7471A	PPB	07 Sep 2017 16:16:06	U	SMPL
CCV55 - 1	CCV55	21509	5.0474	-	7471A	PPB	07 Sep 2017 16:18:23	U	SMPL
CCB55 - 1	CCB55	-309	-0.0507	-	7471A	PPB	07 Sep 2017 16:20:39	U	SMPL
CRA - 1	CRA	704	0.1860	-	7471A	PPB	07 Sep 2017 16:22:54	U	SMPL
HighStd - 1	HighStd	42812	10.0251	-	7471A	PPB	07 Sep 2017 16:25:10	U	SMPL
ChkStd - 1	ChkStd	28577	6.6989	-	7471A	PPB	07 Sep 2017 16:27:26	U	SMPL
PB102109BL - 1	PBS	-385	-0.0684	-	7471A	PPB	07 Sep 2017 16:29:44	U	SMPL
PB102109BS - 1	LCSS	15232	3.5807	-	7471A	PPB	07 Sep 2017 16:32:02	U	SMPL
I5105-01 - 1	OK-02-090517	1153	0.2910	-	7471A	PPB	07 Sep 2017 16:34:19	U	SMPL
I5110-01 - 1	SP-1	6091	1.4448	-	7471A	PPB	07 Sep 2017 16:36:36	U	SMPL
I5110-04 - 1	SP-2	4193	1.0013	-	7471A	PPB	07 Sep 2017 16:38:55	U	SMPL
I5113-01 - 1	RL-1-0-12.25	2047	0.4999	-	7471A	PPB	07 Sep 2017 16:46:02	U	SMPL
I5113-01DUP - 1	RL-1-0-12.25DUP	1980	0.4842	-	7471A	PPB	07 Sep 2017 16:48:17	U	SMPL
CCV56 - 1	CCV56	21317	5.0025	-	7471A	PPB	07 Sep 2017 16:50:33	U	SMPL
CCB56 - 1	CCB56	-285	-0.0450	-	7471A	PPB	07 Sep 2017 16:52:49	U	SMPL
I5113-01MS - 1	RL-1-0-12.25MS	18213	4.2772	-	7471A	PPB	07 Sep 2017 16:55:05	U	SMPL
I5113-01MSD - 1	RL-1-0-12.25MSD	18340	4.3069	-	7471A	PPB	07 Sep 2017 16:57:21	U	SMPL
I5113-03 - 1	RL-1-12.25-24.5	23391	5.4871	-	7471A	PPB	07 Sep 2017 16:59:37	U	SMPL
I5113-05 - 1	RL-3-0-11.75	16632	3.9078	-	7471A	PPB	07 Sep 2017 17:01:54	U	SMPL
I5113-07 - 1	RL-3-11.75-23.5	22129	5.1922	-	7471A	PPB	07 Sep 2017 17:04:11	U	SMPL
I5113-01LX5 - 1		94	0.0435	-	7471A	PPB	07 Sep 2017 17:06:29	U	SMPL
I5113-01A - 1		19132	4.4920	-	7471A	PPB	07 Sep 2017 17:08:46	U	SMPL
CCV57 - 1	CCV57	21449	5.0334	-	7471A	PPB	07 Sep 2017 17:11:03	U	SMPL
CCB57 - 1	CCB57	-265	-0.0404	-	7471A	PPB	07 Sep 2017 17:13:21	U	SMPL