



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB89933

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP57663
COD calibration std. 100 ppm	WP57660
COD calibration std. 10 ppm	WP57659
COD calibration std. 150 ppm	WP57661
COD calibration std. 50 ppm	WP57662
COD calibration std. 0 ppm	WP57658
COD ICV-LCS std, 50ppm	WP57767
COD CCV std, 50ppm	WP57768
COD Digestion Vials Low Level 0-150Mg/L	W2217

Temp In (C): <u>150</u>	Date In: <u>09/07/2017</u>	Time In: <u>12:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>09/07/2017</u>	Time Out: <u>14:01</u>

Intercept: -0.4307

Slope: 1.0166

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	09/01/2017	14:01
2	CAL2	10	1	Water	10.000	09/01/2017	14:01
3	CAL3	50	1	Water	49.000	09/01/2017	14:02
4	CAL4	100	1	Water	102.000	09/01/2017	14:02
5	CAL5	150	1	Water	152.000	09/01/2017	14:03



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB89933

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.607	09/01/2017	14:03
2	ICB				1	Water	0.000	0.424	09/01/2017	14:04
3	CCV1	50			1	Water	49.000	48.624	09/07/2017	15:33
4	CCB1				1	Water	0.000	0.424	09/07/2017	15:33
5	LB89933BL				1	Water	0.000	0.424	09/07/2017	15:34
6	LB89933BS	50			1	Water	50.000	49.607	09/07/2017	15:34
7	I5109-01				1	Water	44.000	43.705	09/07/2017	15:35
8	I5116-02				1	Water	0.000	0.424	09/07/2017	15:35
9	I5116-02DUP				1	Water	0.000	0.424	09/07/2017	15:36
10	I5116-02MS	50			1	Water	51.000	50.591	09/07/2017	15:36
11	I5116-02MSD	50			1	Water	50.000	49.607	09/07/2017	15:37
12	CCV2	50			1	Water	51.000	50.591	09/07/2017	15:37
13	CCB2				1	Water	0.000	0.424	09/07/2017	15:38

2B 89933

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD090717

WorkList ID : 102625

Date : 9/7/2017 10:47:42 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/08/2017	Water	I5109-01	COD	Conc H2SO4 tc	PSEG01	O52	DRUM-1-2-COMP	09/05/2017	SM5220 D
	Water	I5116-02	COD	Conc H2SO4 tc	SHAW22	D11	OUTFALL001-170905	09/05/2017	SM5220 D

Date/Time

Received by:

Relinquished by:

09/07/17 11:05

IB

CP

Date/Time

Received by:

Relinquished by:

09/07/17 12:21

CP

IB