

LB89937

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

9/7/2017 17:30

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.069	0.0	0.116	
ICB1	0.005	0.0	0.002	
CCV1	1.051	0.0	0.114	
CCB1	0.006	0.0	0.003	
PB102134BL	0.006	0.0	0.003	
PB102134BS	1.054	0.0	0.114	
I5134-01	0.006	0.0	0.003	
I5134-01DUP	0.006	0.0	0.003	
I5134-01MS	0.756	0.0	0.082	
I5134-01MS	0.750	0.0	0.082	
CCV2	1.031	0.0	0.112	
CCB2	0.005	0.0	0.002	

N 12
Mean 0.479
SD 0.5049
CV% 105.45

Aquakem v. 7.2AQ1

Results from time period:

Thu Sep 07 16:51:20 2017

Thu Sep 07 17:16:14 2017

L B89937

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0527	mg/l	9/7/2017 10:57:22	
0.1PPM	A	Phenolics-	P	0.1058	mg/l	9/7/2017 10:57:23	
0.25PPM	A	Phenolics-	P	0.2487	mg/l	9/7/2017 10:57:24	
0.50PPM	A	Phenolics-	P	0.4985	mg/l	9/7/2017 10:57:25	
1.0PPM	A	Phenolics-	P	0.9883	mg/l	9/7/2017 10:57:26	
2.0PPM	A	Phenolics-	P	2.006	mg/l	9/7/2017 10:57:27	
ICV1	S	Phenolics-	P	1.0694	mg/l	9/7/2017 16:51:21	
ICB1	S	Phenolics-	P	0.0055	mg/l	9/7/2017 16:51:22	
CCV1	S	Phenolics-	P	1.0506	mg/l	9/7/2017 16:51:25	
CCB1	S	Phenolics-	P	0.006	mg/l	9/7/2017 16:51:26	
PB102134BL	S	Phenolics-	P	0.0065	mg/l	9/7/2017 16:51:28	
PB102134BS	S	Phenolics-	P	1.054	mg/l	9/7/2017 16:51:31	
I5134-01	S	Phenolics-	P	0.0059	mg/l	9/7/2017 16:58:06	
I5134-01DUP	S	Phenolics-	P	0.0057	mg/l	9/7/2017 16:58:07	
I5134-01MS	S	Phenolics-	P	0.7562	mg/l	9/7/2017 16:58:08	
I5134-01MSD	S	Phenolics-	P	0.7505	mg/l	9/7/2017 17:16:11	
CCV2	S	Phenolics-	P	1.0308	mg/l	9/7/2017 17:16:13	
CCB2	S	Phenolics-	P	0.005	mg/l	9/7/2017 17:16:14	

LB89937

=====
Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

9/7/2017 9:11

Test Phenolics-


Accepted 9/7/2017 9:11

Factor 7.861

Bias 0.002

Coeff. of det. 0.999701

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0575	0.0500	
2	Phenol-2	0.015	0.1034	0.1000	
3	Phenol-2	0.034	0.2532	0.2500	
4	Phenol-2	0.065	0.4987	0.5000	
5	Phenol-2	0.126	0.9753	1.0000	
6	Phenol-2	0.258	2.0119	2.0000	