



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89940

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/07/2017	16:30
2	CAL2	1	Water			7.01	09/07/2017	16:31
3	CAL3	1	Water			10.04	09/07/2017	16:33
4	ICV	1	Water			7.00	09/07/2017	16:35
5	CCV1	1	Water			2.01	09/07/2017	16:36
6	I5137-02	1	Solid	20.02	20	8.07	09/07/2017	16:37
7	I5158-03	1	Solid	20.03	20	7.63	09/07/2017	16:38
8	I5158-03DUP	1	Solid	20.04	20	7.64	09/07/2017	16:39
9	CCV2	1	Water			12.01	09/07/2017	16:40

WORKLIST(Hardcopy Internal Chain)

1589940

WorkList Name : CORROSIONITY I5137

WorkList ID : 102642

Date : 9/7/2017 12:59:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2017	Solid	I5137-02	Corrosivity	Cool 4 deg C	PSEG01	P31	TP-1-COMP	09/06/2017	9045C
09/13/2017	Solid	I5158-03	Corrosivity	Cool 4 deg C	PSEG05	D52	BU-02-090717	09/07/2017	9045C

Date/Time 09-07-17 12:00 PM
 Received by: JD
 Relinquished by: CP

Date/Time 09-07-17 6:20 PM
 Received by: CP
 Relinquished by: JD