



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB89943
Analysis Date: 09/08/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/08/2017
Extraction IN Time: 09:30
Extraction OUT Time: 10:20

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB89943BL	LB89943BL	WATER	2	1000	100	3.0179	3.0179	3.02	3.0179	3.0179	0	0
2	LB89943BS	LB89943BS	WATER	2	1000	100	3.1562	3.1562	3.01	3.1751	3.1751	0.0189	18.9
3	I5130-02	GRAB	WATER	2	1000	100	3.0298	3.0298	3.04	3.0886	3.0886	0.0588	58.8
4	I5152-03	DRAIN-WATER-TANK-1	WATER	2	500	100	2.9277	2.9277	3.02	2.9311	2.9311	0.0034	6.8
5	I5152-03MS	DRAIN-WATER-TANK-1	WATER	2	500	100	2.7405	2.7405	3.04	2.7538	2.7538	0.0133	26.6
6	I5152-03MSD	DRAIN-WATER-TANK-1	WATER	2	500	100	2.7133	2.7133	3.03	2.7261	2.7261	0.0128	25.6
7	I5157-01	DSN001	WATER	2	970	100	2.9289	2.9277	3.01	2.9297	2.9297	0.0008	0.82



QC Batch# LB89943 Test: TPH Analysis Date: 09/08/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2272
pH Paper 0-14	W2248
Sodium Sulfate	EP1774
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP56784
LCSWD		
MS/MSD	10.00 ML	WP54439

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 11:30
1.0000 gram Balance: 1.0005 In Time: 11:00
Bal Check Time: 09:40

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 13:10
1.0000 gram Balance: 1.0003 Out Time: 12:30
Bal Check Time: 13:55

2589943

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I5149

WorkList ID : 102682

Date : 9/8/2017 9:11:02 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5130-02 ✓	TPH	Conc H2SO4 4c	ARAM01	D11	GRAB	09/06/2017	1664A
	Water	I5152-03 D.B.C.	TPH	Conc H2SO4 4c	MAJO01	D11	DRAIN-WATER-TANK-1	09/07/2017	1664A
	Water	I5157-01 -1	TPH	Conc H2SO4 4c	NATI04	D41	DSN001	09/07/2017	1664A

Date/Time 09.08.17 9:11 AM
Received by: JP
Relinquished by: CB

Date/Time 09.08.17 11:50 AM
Received by: CB
Relinquished by: JP