


7470A

LB 89947
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35201



A= 0.0000e+000

B= 2.8375e-004

C= 7.7268e-003

Rho= 0.9999887

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	-88	0.000	-88				
0.2	0.200	0.227	0.027	774	0.0 %	774				
2.5	2.500	2.477	-0.023	8701	0.0 %	8701				
5.0	5.000	5.015	0.015	17648	0.0 %	17648				
7.5	7.500	7.502	0.002	26411	0.0 %	26411				
10.0	10.000	9.996	-0.004	35201	0.0 %	35201				

LB89947 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 08 Sep 2017 11:17:38					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	-88	-	0.0000	7470A	PPB	08 Sep 2017 11:19:11	S	Std
0.2 - 1	30-2	774	-	0.2000	7470A	PPB	08 Sep 2017 11:21:19	S	Std
2.5 - 1	32.5	8701	-	2.5000	7470A	PPB	08 Sep 2017 11:23:28	S	Std
5.0 - 1	35	17648	-	5.0000	7470A	PPB	08 Sep 2017 11:25:36	S	Std
7.5 - 1	37.5	26411	-	7.5000	7470A	PPB	08 Sep 2017 11:27:44	S	Std
10.0 - 1	40	35201	-	10.0000	7470A	PPB	08 Sep 2017 11:29:51	S	Std
ICV46 - 1	ICV46	13479	3.8324	-7470A	PPB	PPB	08 Sep 2017 11:32:44	U	SMPL
ICB46 - 1	ICB46	-186	-0.0451	-7470A	PPB	PPB	08 Sep 2017 11:34:51	U	SMPL
CCV58 - 1	CCV58	17832	5.0676	-7470A	PPB	PPB	08 Sep 2017 11:36:57	U	SMPL
CCB58 - 1	CCB58	-238	-0.0598	-7470A	PPB	PPB	08 Sep 2017 11:39:04	U	SMPL
CRA - 1	CRA	635	0.1879	-7470A	PPB	PPB	08 Sep 2017 11:41:10	U	SMPL
HighStd - 1	HighStd	35750	10.1518	-7470A	PPB	PPB	08 Sep 2017 11:43:17	U	SMPL
ChkStd - 1	ChkStd	25507	7.2454	-7470A	PPB	PPB	08 Sep 2017 11:45:24	U	SMPL
PB102128BL - 1	PBW	-189	-0.0459	-7470A	PPB	PPB	08 Sep 2017 11:49:56	U	SMPL
PB102128BS - 1	LCSW	13724	3.9019	-7470A	PPB	PPB	08 Sep 2017 11:52:03	U	SMPL
PB102128TB - 1	PB102110TB	-260	-0.0660	-7470A	PPB	PPB	08 Sep 2017 11:54:11	U	SMPL
I5120-03 - 1	B-16(0-10)COMP	-148	-0.0343	-7470A	PPB	PPB	08 Sep 2017 11:56:19	U	SMPL
I5122-03 - 1	SU-04-090617	-191	-0.0465	-7470A	PPB	PPB	08 Sep 2017 11:58:27	U	SMPL
I5126-03 - 1	TR-05-090617	-130	-0.0292	-7470A	PPB	PPB	08 Sep 2017 12:00:36	U	SMPL
I5133-02 - 1	VNJ-204	-113	-0.0243	-7470A	PPB	PPB	08 Sep 2017 12:02:43	U	SMPL
CCV59 - 1	CCV59	17816	5.0630	-7470A	PPB	PPB	08 Sep 2017 12:04:49	U	SMPL
CCB59 - 1	CCB59	-246	-0.0621	-7470A	PPB	PPB	08 Sep 2017 12:06:56	U	SMPL
I5135-02 - 1	EPI-280-283	-170	-0.0405	-7470A	PPB	PPB	08 Sep 2017 12:09:03	U	SMPL
I5136-02 - 1	PAINT-CHIPS	-90	-0.0178	-7470A	PPB	PPB	08 Sep 2017 12:11:10	U	SMPL
I5136-04 - 1	BRICK-WALL	-124	-0.0275	-7470A	PPB	PPB	08 Sep 2017 12:13:17	U	SMPL
I5136-04DUP - 1	BRICK-WALLDUP	-125	-0.0277	-7470A	PPB	PPB	08 Sep 2017 12:15:24	U	SMPL
I5136-04MS - 1	BRICK-WALLMS	13253	3.7683	-7470A	PPB	PPB	08 Sep 2017 12:17:31	U	SMPL
I5136-04MSD - 1	BRICK-WALLMSD	13086	3.7209	-7470A	PPB	PPB	08 Sep 2017 12:19:38	U	SMPL
I5136-04LX5 - 1		-289	-0.0743	-7470A	PPB	PPB	08 Sep 2017 12:21:46	U	SMPL
I5136-04A - 1		14416	4.0983	-7470A	PPB	PPB	08 Sep 2017 12:23:55	U	SMPL
CCV60 - 1	CCV60	17880	5.0812	-7470A	PPB	PPB	08 Sep 2017 12:26:03	U	SMPL
CCB60 - 1	CCB60	-242	-0.0609	-7470A	PPB	PPB	08 Sep 2017 12:28:10	U	SMPL