

245.1


LB89950

INSTRUMENT ID: CV1

Linear

 μ Abs.:

31592



A= 0.0000e+000

B= 3.1679e-004

C= 3.7116e-003

Rho= 0.9999485

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-122	0.000	-122				
0.2	0.200	0.216	0.016	671	0.0 %	671				
2.5	2.500	2.507	0.007	7903	0.0 %	7903				
5.0	5.000	5.058	0.058	15954	0.0 %	15954				
7.5	7.500	7.442	-0.058	23479	0.0 %	23479				
10.0	10.000	10.012	0.012	31592	0.0 %	31592				

LB89950 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Sep 2017 12:57:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-122	-	0.0000	245.1	PPB	08 Sep 2017 12:59:41	S	Std
0.2 - 1	50.2	671	-	0.2000	245.1	PPB	08 Sep 2017 13:01:49	S	Std
2.5 - 1	52.5	7903	-	2.5000	245.1	PPB	08 Sep 2017 13:03:58	S	Std
5.0 - 1	55	15954	-	5.0000	245.1	PPB	08 Sep 2017 13:06:07	S	Std
7.5 - 1	57.5	23479	-	7.5000	245.1	PPB	08 Sep 2017 13:08:14	S	Std
10.0 - 1	50	31592	-	10.0000	245.1	PPB	08 Sep 2017 13:10:54	S	Std
ICV47 - 1	ICV47	12270	3.8908	-245.1	PPB	PPB	08 Sep 2017 13:13:22	U	SMPL
ICB47 - 1	ICB47	-314	-0.0958	-245.1	PPB	PPB	08 Sep 2017 13:15:28	U	SMPL
CCV61 - 1	CCV61	15865	5.0296	-245.1	PPB	PPB	08 Sep 2017 13:17:35	U	SMPL
CCB61 - 1	CCB61	-315	-0.0961	-245.1	PPB	PPB	08 Sep 2017 13:19:42	U	SMPL
CRA - 1	CRA	654	0.2109	-245.1	PPB	PPB	08 Sep 2017 13:21:49	U	SMPL
HighStd - 1	HighStd	31471	9.9735	-245.1	PPB	PPB	08 Sep 2017 13:23:55	U	SMPL
ChkStd - 1	ChkStd	22117	7.0102	-245.1	PPB	PPB	08 Sep 2017 13:26:02	U	SMPL
PB102129BL - 1	PBW	-289	-0.0878	-245.1	PPB	PPB	08 Sep 2017 13:28:11	U	SMPL
PB102129BS - 1	LCSW	13616	4.3172	-245.1	PPB	PPB	08 Sep 2017 13:30:19	U	SMPL
I5124-02 - 1	EFF	-188	-0.0558	-245.1	PPB	PPB	08 Sep 2017 13:32:27	U	SMPL
I5124-02DUP - 1	EFFDUP	-35	-0.0074	-245.1	PPB	PPB	08 Sep 2017 13:34:34	U	SMPL
I5124-02MS - 1	EFFMS	10152	3.2198	-245.1	PPB	PPB	08 Sep 2017 13:36:43	U	SMPL
I5124-02MSD - 1	EFFMSD	10841	3.4381	-245.1	PPB	PPB	08 Sep 2017 13:38:51	U	SMPL
I5130-01 - 1	COMP	-97	-0.0270	-245.1	PPB	PPB	08 Sep 2017 13:40:59	U	SMPL
CCV62 - 1	CCV62	15834	5.0198	-245.1	PPB	PPB	08 Sep 2017 13:43:07	U	SMPL
CCB62 - 1	CCB62	-322	-0.0983	-245.1	PPB	PPB	08 Sep 2017 13:45:14	U	SMPL
I5131-01 - 1	090517-TIDO-F-D-B	-208	-0.0622	-245.1	PPB	PPB	08 Sep 2017 13:47:21	U	SMPL
I5131-02 - 1	090517-TIDO-F-D-S	-287	-0.0872	-245.1	PPB	PPB	08 Sep 2017 13:49:28	U	SMPL
I5131-04 - 1	090517-TIDO-F-U-B	-259	-0.0783	-245.1	PPB	PPB	08 Sep 2017 13:51:35	U	SMPL
I5131-05 - 1	090517-TIDO-F-U-S	-240	-0.0723	-245.1	PPB	PPB	08 Sep 2017 13:53:42	U	SMPL
I5132-01 - 1	090617-TIDO-F-D-B	-272	-0.0825	-245.1	PPB	PPB	08 Sep 2017 13:55:49	U	SMPL
I5132-02 - 1	090617-TIDO-F-D-S	-297	-0.0904	-245.1	PPB	PPB	08 Sep 2017 13:57:57	U	SMPL
I5132-03 - 1	090617-TIDO-F-U-B	-285	-0.0866	-245.1	PPB	PPB	08 Sep 2017 14:00:05	U	SMPL
I5132-04 - 1	090617-TIDO-F-U-S	-309	-0.0942	-245.1	PPB	PPB	08 Sep 2017 14:02:13	U	SMPL
I5124-02LX5 - 1		-169	-0.0498	-245.1	PPB	PPB	08 Sep 2017 14:04:22	U	SMPL
I5124-02A - 1		11079	3.5135	-245.1	PPB	PPB	08 Sep 2017 14:06:28	U	SMPL
CCV63 - 1	CCV63	15987	5.0683	-245.1	PPB	PPB	08 Sep 2017 14:08:35	U	SMPL
CCB63 - 1	CCB63	-310	-0.0945	-245.1	PPB	PPB	08 Sep 2017 14:10:43	U	SMPL