



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89952

Temp.: 17.8 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/08/2017	13:40
2	CAL2	1	Water			7.00	09/08/2017	13:41
3	CAL3	1	Water			10.03	09/08/2017	13:42
4	ICV	1	Water			6.99	09/08/2017	13:43
5	CCV1	1	Water			2.01	09/08/2017	13:45
6	I5124-01	1	Water			6.98	09/08/2017	13:46
7	I5164-01	1	Water			6.38	09/08/2017	13:47
8	I5164-01DUP	1	Water			6.39	09/08/2017	13:48
9	CCV2	1	Water			12.01	09/08/2017	13:50

1389952

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5164 WorkList ID : 102691 Date : 9/8/2017 1:30:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/06/2017	Water	I5124-01-13	pH	Cool 4 deg C	ARDM01	D32	EFF	09/06/2017	SM 4500-PH B
	Water	I5164-01-D	pH	Cool 4 deg C	PARS02	D41	SES-MS-PH4-COA9-17	09/07/2017	SM 4500-PH B

Date/Time 09/08/17 1:30 PM
Received by: JD
Relinquished by: _____

Date/Time 09/08/17 2:00 PM
Received by: CR
Relinquished by: JD