

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Dhvani

Date: 09/08/2017

DATE & TIME IN: 09/08/2017 10:20

OVEN TEMP: 103 °C

Run Number: LB89958

DATE & TIME OUT: 09/08/2017 14:30

OVEN TEMP: 104 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB89958BL	LB89958BL	1.4408	1.4408	100	1.4408	1.4408	1.4408	0	0
2	LB89958BS	LB89958BS	1.4218	1.4218	100	1.4756	1.4756	1.4756	0.0538	538
3	I5143-01	TREATEDEFFLUENT-MW-2	1.42	1.42	100	1.4234	1.4234	1.4234	0.0034	34
4	I5143-01DUP	TREATEDEFFLUENT-MW-2DUP	1.4013	1.4013	100	1.4046	1.4046	1.4046	0.0033	33

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB89958

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 090717TSS-1 WorkList ID : 102646 Date : 9/7/2017 1:48:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5143-01 ✓	TSS	Cool 4 deg C	TULL02	D11	TREATEDEFFLUENT-MW-2	09/06/2017	SM2540 D

Date/Time 09/08/17 9:30
 Received by: DP
 Relinquished by: DP

Date/Time 09/08/17 10:30
 Received by: DP
 Relinquished by: DP