



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB89971

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP57663
COD calibration std. 100 ppm	WP57660
COD calibration std. 10 ppm	WP57659
COD calibration std. 150 ppm	WP57661
COD calibration std. 50 ppm	WP57662
COD calibration std. 0 ppm	WP57658
COD ICV-LCS std, 50ppm	WP57851
COD CCV std, 50ppm	WP57852
COD Digestion Vials Low Level 0-150Mg/L	W2217

Temp In (C): <u>150</u>	Date In: <u>09/11/2017</u>	Time In: <u>09:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>09/11/2017</u>	Time Out: <u>11:02</u>

Intercept: -0.4307

Slope: 1.0166

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	09/01/2017	14:01
2	CAL2	10	1	Water	10.000	09/01/2017	14:01
3	CAL3	50	1	Water	49.000	09/01/2017	14:02
4	CAL4	100	1	Water	102.000	09/01/2017	14:02
5	CAL5	150	1	Water	152.000	09/01/2017	14:03



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB89971

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.607	09/01/2017	14:03
2	ICB				1	Water	0.000	0.424	09/01/2017	14:04
3	CCV1	50			1	Water	52.000	51.575	09/11/2017	12:05
4	CCB1				1	Water	0.000	0.424	09/11/2017	12:05
5	LB89971BL				1	Water	0.000	0.424	09/11/2017	12:06
6	LB89971BS	50			1	Water	51.000	50.591	09/11/2017	12:06
7	I5157-01				1	Water	15.000	15.179	09/11/2017	12:07
8	I5157-01DUP				1	Water	14.000	14.195	09/11/2017	12:07
9	I5157-01MS	50			1	Water	68.000	67.313	09/11/2017	12:08
10	I5157-01MSD	50			1	Water	67.000	66.330	09/11/2017	12:08
11	I5164-01				1	Water	6.000	6.326	09/11/2017	12:09
12	CCV2	50			1	Water	52.000	51.575	09/11/2017	12:09
13	CCB2				1	Water	0.000	0.424	09/11/2017	12:10

WORKLIST(Hardcopy Internal Chain)

2389971

WorkList Name : COD090817

WorkList ID : 102700

Date : 9/8/2017 4:19:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5164-01	COD	Conc H2SO4 4c	PARS02	D41	SES-MS-PH4-COA9-17	09/07/2017	SM5220 D
	Water	I5157-01	COD	Conc H2SO4 4c	NAT104	D41	DSN001	09/07/2017	SM5220 D

Date/Time 09/11/17 8:45
 Received by: IB
 Relinquished by: CP

Date/Time 09/11/17 10:00
 Received by: CP
 Relinquished by: IB