

ISM02.4


LB 89973

INSTRUMENT ID : CVI

Linear

 μ Abs.:

40523



A= 0.0000e+000

B= 2.4986e-004

C= -1.1314e-001

Rho= 0.9999588

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0/0
0.00	0.000	-0.020	-0.020	374	0.000	374					-
0.2	0.200	0.180	-0.020	1172	0.0 %	1172					-10
2.5	2.500	2.561	0.061	10701	0.0 %	10701					2
5.0	5.000	5.012	0.012	20511	0.0 %	20511					0
7.5	7.500	7.456	-0.044	30292	0.0 %	30292					-1
10.0	10.000	10.012	0.012	40523	0.0 %	40523					0

NB
9/8/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	374	-		0 ISM02.4	PPB	9/8/2017 19:44	Std
	0.2 S0.2	1172	-		0.2 ISM02.4	PPB	9/8/2017 19:46	Std
	2.5 S2.5	10701	-		2.5 ISM02.4	PPB	9/8/2017 19:48	Std
	5 S5	20511	-		5 ISM02.4	PPB	9/8/2017 19:51	Std
	7.5 S7.5	30292	-		7.5 ISM02.4	PPB	9/8/2017 19:53	Std
	10 S10	40523	-		10 ISM02.4	PPB	9/8/2017 19:55	Std
ICV	ICV	15748	3.8217 -		ISM02.4	PPB	9/8/2017 19:58	SMPL
ICB	ICB	201	-0.0629 -		ISM02.4	PPB	9/8/2017 20:00	SMPL
CCV074	CCV074	20381	4.9793 -		ISM02.4	PPB	9/8/2017 20:03	SMPL
CCB074	CCB074	190	-0.0657 -		ISM02.4	PPB	9/8/2017 20:05	SMPL
PB102164BL	PBW003	234	-0.0547 -		ISM02.4	PPB	9/8/2017 20:07	SMPL
I5150-07	MX2B01	35388	8.7289 -		ISM02.4	PPB	9/8/2017 20:12	SMPL
I5150-08	MX2B01D	35831	8.8396 -		ISM02.4	PPB	9/8/2017 20:14	SMPL
I5150-09	MX2B01S	39442	9.7419 -		ISM02.4	PPB	9/8/2017 20:16	SMPL
CCV075	CCV075	20213	4.9373 -		ISM02.4	PPB	9/8/2017 20:19	SMPL
CCB075	CCB075	149	-0.0759 -		ISM02.4	PPB	9/8/2017 20:21	SMPL