

ISM02.4


LB 89974

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44789



A= 0.0000e+000

B= 2.2549e-004

C= -8.0032e-002

Rho= 0.9999788

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0D
0.00	0.000	-0.009	-0.009	317	0.000	317					-
0.2	0.200	0.199	-0.001	1238	0.0 %	1238					-1
2.5	2.500	2.540	0.040	11621	0.0 %	11621					2
5.0	5.000	4.968	-0.032	22388	0.0 %	22388					-1
7.5	7.500	7.481	-0.019	33531	0.0 %	33531					0
10.0	10.000	10.020	0.020	44789	0.0 %	44789					0

LB
9/8/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	317	-		0 ISM02.4	PPB	9/8/2017 20:30	Std
	0.2 S0.2	1238	-		0.2 ISM02.4	PPB	9/8/2017 20:33	Std
	2.5 S2.5	11621	-		2.5 ISM02.4	PPB	9/8/2017 20:35	Std
	5 S5	22388	-		5 ISM02.4	PPB	9/8/2017 20:37	Std
	7.5 S7.5	33531	-		7.5 ISM02.4	PPB	9/8/2017 20:40	Std
	10 S10	44789	-		10 ISM02.4	PPB	9/8/2017 20:42	Std
ICV	ICV	17725	3.9169 -		ISM02.4	PPB	9/8/2017 20:45	SMPL
ICB	ICB	273	-0.0185 -		ISM02.4	PPB	9/8/2017 20:47	SMPL
CCV076	CCV076	22403	4.9717 -		ISM02.4	PPB	9/8/2017 20:49	SMPL
CCB076	CCB076	305	-0.0113 -		ISM02.4	PPB	9/8/2017 20:51	SMPL
PB102152BL	PBS004	429	0.0167 -		ISM02.4	PPB	9/8/2017 20:54	SMPL
I5150-19	MX2B05	30625	6.8257 -		ISM02.4	PPB	9/8/2017 20:58	SMPL
I5150-20	MX2B05D	32606	7.2724 -		ISM02.4	PPB	9/8/2017 21:01	SMPL
I5150-21	MX2B05S	41971	9.3842 -		ISM02.4	PPB	9/8/2017 21:03	SMPL
CCV077	CCV077	21753	4.8251 -		ISM02.4	PPB	9/8/2017 21:05	SMPL
CCB077	CCB077	163	-0.0433 -		ISM02.4	PPB	9/8/2017 21:07	SMPL