



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB89981
Analysis Date: 09/11/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/11/2017
Extraction IN Time: 08:20
Extraction OUT Time: 09:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB89981BL	LB89981BL	WATER	2	1000	100	3.0666	3.0666	3.02	3.0666	3.0666	0	0
2	LB89981BS	LB89981BS	WATER	2	1000	100	3.1852	3.1852	3.04	3.2041	3.2041	0.0189	18.9
3	I5190-01	EFF-WASTE-WATER	WATER	2	980	100	3.0274	3.0274	3.02	3.0605	3.0605	0.0331	33.78
4	I5190-01MS	EFF-WASTE-WATER	WATER	2	980	100	3.0411	3.0411	3.03	3.0949	3.0949	0.0538	54.9



QC Batch# LB89981 Test: TPH Analysis Date: 09/11/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2272
pH Paper 0-14	W2248
Sodium Sulfate	EP1774
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP56784
LCSWD		
MS/MSD	10.00 ML	WP54439

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0019 In OVEN TEMP : 71 Dessicator Time In : 10:15
1.0000 gram Balance: 1.0003 In Time: 09:40
Bal Check Time: 08:40

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 70 Dessicator Time Out: 12:00
1.0000 gram Balance: 1.0005 Out Time: 11:15
Bal Check Time: 12:30

1389981

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I5190

WorkList ID : 102741

Date : 9/11/2017 7:57:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5190-01	TPH	Conc H2SO4 to	ARDM01	D22	EFF-WASTE-WATER	09/06/2017	1664A

-B, C,

Date/Time 09-11-17 8:00 PM
 Received by: JP
 Relinquished by: CA

Date/Time 09-11-17 10:00 PM
 Received by: CD
 Relinquished by: JP