

7471A


LB 89982

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43131



A= 0.0000e+000

B= 2.3231e-004

C= 3.7814e-002

Rho= 0.9998688

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.038	0.038	-1	0.000	-1				
0.2	0.200	0.190	-0.010	655	0.0 %	655				
2.5	2.500	2.550	0.050	10814	0.0 %	10814				
5.0	5.000	4.883	-0.117	20855	0.0 %	20855				
7.5	7.500	7.482	-0.018	32044	0.0 %	32044	0			
10.0	10.000	10.058	0.058	43131	0.0 %	43131				

LB89982 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Sep 2017 09:43:39

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	58	-1	-	0.00007471A	PPB	11 Sep 2017 09:51:50	S	Std
0.2 - 1	50.2	655	-	0.20007471A	PPB	11 Sep 2017 09:54:08	S	Std
2.5 - 1	52.5	10814	-	2.50007471A	PPB	11 Sep 2017 09:56:25	S	Std
5.0 - 1	55	20855	-	5.00007471A	PPB	11 Sep 2017 09:58:43	S	Std
7.5 - 1	57.5	32044	-	7.50007471A	PPB	11 Sep 2017 10:01:00	S	Std
10.0 - 1	510	43131	-	10.00007471A	PPB	11 Sep 2017 10:07:49	S	Std
ICV49 - 1	ICV49	16713	3.9205	-7471A	PPB	11 Sep 2017 10:10:33	U	SMPL
ICB49 - 1	ICB49	-164	-0.0003	-7471A	PPB	11 Sep 2017 10:12:49	U	SMPL
CCV78 - 1	CCV78	21557	5.0458	-7471A	PPB	11 Sep 2017 10:15:05	U	SMPL
CCB78 - 1	CCB78	-195	-0.0075	-7471A	PPB	11 Sep 2017 10:17:20	U	SMPL
CRA - 1	CRA	568	0.1698	-7471A	PPB	11 Sep 2017 10:19:36	U	SMPL
HighStd - 1	HighStd	40282	9.3958	-7471A	PPB	11 Sep 2017 10:21:52	U	SMPL
ChkStd - 1	ChkStd	27875	6.5135	-7471A	PPB	11 Sep 2017 10:24:08	U	SMPL
PB102178BL - 1	PBS	-290	-0.0296	-7471A	PPB	11 Sep 2017 10:26:25	U	SMPL
PB102178BS - 1	LCSS	15339	3.6013	-7471A	PPB	11 Sep 2017 10:28:43	U	SMPL
I5171-01 - 1	RT-2342-2669-3025-3091-3873	1844	0.4662	-7471A	PPB	11 Sep 2017 10:31:00	U	SMPL
I5178-02 - 1	TEC-COMP-0-1	2889	0.7090	-7471A	PPB	11 Sep 2017 10:33:17	U	SMPL
I5178-03 - 1	TEC-COMP-1-3	12425	2.9243	-7471A	PPB	11 Sep 2017 10:35:34	U	SMPL
I5178-04 - 1	TEC-COMP-3-5	1866	0.4713	-7471A	PPB	11 Sep 2017 10:37:52	U	SMPL
CCV79 - 1	CCV79	21384	5.0056	-7471A	PPB	11 Sep 2017 10:40:08	U	SMPL
CCB79 - 1	CCB79	-184	-0.0049	-7471A	PPB	11 Sep 2017 10:42:23	U	SMPL
I5178-05 - 1	TEC-COMP-5-7	989	0.2676	-7471A	PPB	11 Sep 2017 10:44:40	U	SMPL
I5178-05DUP - 1	TEC-COMP-5-7DUP	823	0.2290	-7471A	PPB	11 Sep 2017 10:46:56	U	SMPL
I5178-05MS - 1	TEC-COMP-5-7MS	16049	3.7662	-7471A	PPB	11 Sep 2017 10:49:12	U	SMPL
I5178-05MSD - 1	TEC-COMP-5-7MSD	15925	3.7374	-7471A	PPB	11 Sep 2017 10:51:28	U	SMPL
I5178-06 - 1	TEC-COMP-7-9	3030	0.7417	-7471A	PPB	11 Sep 2017 10:53:44	U	SMPL
I5178-07 - 1	TEC-COMP-9-11	1161	0.3075	-7471A	PPB	11 Sep 2017 10:56:00	U	SMPL
I5178-08 - 1	TEC-COMP-11-12	103	0.0617	-7471A	PPB	11 Sep 2017 10:58:17	U	SMPL
I5184-01 - 1	OK-02-090817	1704	0.4337	-7471A	PPB	11 Sep 2017 11:00:35	U	SMPL
I5186-01 - 1	284-GRAB	6626	1.5771	-7471A	PPB	11 Sep 2017 11:16:04	U	SMPL
CCV80 - 1	CCV80	21168	4.9554	-7471A	PPB	11 Sep 2017 11:18:21	U	SMPL
CCB80 - 1	CCB80	-244	-0.0189	-7471A	PPB	11 Sep 2017 11:20:38	U	SMPL
I5178-05LX5 - 1		110	0.0634	-7471A	PPB	11 Sep 2017 11:22:56	U	SMPL
I5178-05A - 1		18120	4.2473	-7471A	PPB	11 Sep 2017 11:25:13	U	SMPL
CCV81 - 1	CCV81	21320	4.9907	-7471A	PPB	11 Sep 2017 11:27:29	U	SMPL
CCB81 - 1	CCB81	-223	-0.0140	-7471A	PPB	11 Sep 2017 11:29:45	U	SMPL