


7470A

LB89985
INSTRUMENT ID: CV2

Linear

 μ Abs.:

37010



A= 0.0000e+000

B= 2.7235e-004

C= -2.7989e-003

Rho= 0.999827

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	22	0.000	22				
0.2	0.200	0.206	0.006	765	0.0 %	765				
2.5	2.500	2.537	0.037	9326	0.0 %	9326				
5.0	5.000	4.985	-0.015	18312	0.0 %	18312				
7.5	7.500	7.392	-0.108	27153	0.0 %	27153				
10.0	10.000	10.077	0.077	37010	0.0 %	37010				

LB89985 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 11 Sep 2017 11:48:00				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	50	22	-	0.0000	7470A	PPB	11 Sep 2017	11:50:15	S	Std
0.2 - 1	30.2	765	-	0.2000	7470A	PPB	11 Sep 2017	11:52:23	S	Std
2.5 - 1	32.5	9326	-	2.5000	7470A	PPB	11 Sep 2017	11:54:32	S	Std
5.0 - 1	55	18312	-	5.0000	7470A	PPB	11 Sep 2017	11:56:40	S	Std
7.5 - 1	57.5	27153	-	7.5000	7470A	PPB	11 Sep 2017	11:58:47	S	Std
10.0 - 1	510	37010	-	10.0000	7470A	PPB	11 Sep 2017	12:03:25	S	Std
ICV50 - 1	ICV50	13976	3.8036	-	7470A	PPB	11 Sep 2017	12:05:55	U	SMPL
ICB50 - 1	ICB50	-123	-0.0363	-	7470A	PPB	11 Sep 2017	12:08:02	U	SMPL
CCV82 - 1	CCV82	17878	4.8664	-	7470A	PPB	11 Sep 2017	12:10:09	U	SMPL
CCB82 - 1	CCB82	-155	-0.0450	-	7470A	PPB	11 Sep 2017	12:12:15	U	SMPL
CRA - 1	CRA	735	0.1974	-	7470A	PPB	11 Sep 2017	12:14:23	U	SMPL
HighStd - 1	HighStd	37797	10.2914	-	7470A	PPB	11 Sep 2017	12:16:29	U	SMPL
ChkStd - 1	ChkStd	25264	6.8780	-	7470A	PPB	11 Sep 2017	12:18:36	U	SMPL
PB102176BL - 1	PBW	-228	-0.0649	-	7470A	PPB	11 Sep 2017	12:20:45	U	SMPL
PB102176BS - 1	LCSW	12636	3.4387	-	7470A	PPB	11 Sep 2017	12:22:55	U	SMPL
PB102176TB - 1	PB102132TB	-65	-0.0205	-	7470A	PPB	11 Sep 2017	12:25:02	U	SMPL
I5137-02 - 1	TP-1-COMP	-19	-0.0080	-	7470A	PPB	11 Sep 2017	12:27:11	U	SMPL
I5151-07 - 1	NYSEG-PH4-ESMI2-1	-59	-0.0189	-	7470A	PPB	11 Sep 2017	12:29:19	U	SMPL
I5151-08 - 1	NYSEG-PH4-ESMI2-2	-52	-0.0170	-	7470A	PPB	11 Sep 2017	12:31:28	U	SMPL
I5151-09 - 1	NYSEG-PH4-ESMI2-3	-12	-0.0061	-	7470A	PPB	11 Sep 2017	12:33:35	U	SMPL
CCV83 - 1	CCV83	18338	4.9916	-	7470A	PPB	11 Sep 2017	12:35:42	U	SMPL
CCB83 - 1	CCB83	-127	-0.0374	-	7470A	PPB	11 Sep 2017	12:37:48	U	SMPL
I5151-10 - 1	NYSEG-PH4-ESMI2-4	-21	-0.0085	-	7470A	PPB	11 Sep 2017	12:39:57	U	SMPL
I5151-11 - 1	NYSEG-PH4-ESMI2-5	-30	-0.0110	-	7470A	PPB	11 Sep 2017	12:42:04	U	SMPL
I5151-12 - 1	NYSEG-PH4-ESMI2-6	-69	-0.0216	-	7470A	PPB	11 Sep 2017	12:44:11	U	SMPL
I5158-03 - 1	BU-02-090717	-63	-0.0200	-	7470A	PPB	11 Sep 2017	12:46:18	U	SMPL
I5158-03DUP - 1	BU-02-090717DUP	-25	-0.0096	-	7470A	PPB	11 Sep 2017	12:48:25	U	SMPL
I5158-03MS - 1	BU-02-090717MS	13595	3.6999	-	7470A	PPB	11 Sep 2017	12:50:33	U	SMPL
I5158-03MSD - 1	BU-02-090717MSD	14041	3.8213	-	7470A	PPB	11 Sep 2017	12:52:41	U	SMPL
PB102177BL - 1	PBW	-138	-0.0404	-	7470A	PPB	11 Sep 2017	12:54:50	U	SMPL
PB102177BS - 1	LCSW	13324	3.6261	-	7470A	PPB	11 Sep 2017	12:56:59	U	SMPL
I5164-01 - 1	SES-MS-PH4-COA9-17	-339	-0.0951	-	7470A	PPB	11 Sep 2017	12:59:06	U	SMPL
CCV84 - 1	CCV84	18491	5.0333	-	7470A	PPB	11 Sep 2017	13:01:13	U	SMPL
CCB84 - 1	CCB84	-133	-0.0390	-	7470A	PPB	11 Sep 2017	13:03:20	U	SMPL
I5164-01DUP - 1	SES-MS-PH4-COA9-17DUP	-259	-0.0733	-	7470A	PPB	11 Sep 2017	13:05:27	U	SMPL
I5164-01MS - 1	SES-MS-PH4-COA9-17MS	13427	3.6541	-	7470A	PPB	11 Sep 2017	13:07:34	U	SMPL
I5164-01MSD - 1	SES-MS-PH4-COA9-17MSD	13191	3.5898	-	7470A	PPB	11 Sep 2017	13:09:41	U	SMPL
I5167-01 - 1	CPS123061	-154	-0.0447	-	7470A	PPB	11 Sep 2017	13:11:48	U	SMPL
I5167-02 - 1	CPS103061	-212	-0.0605	-	7470A	PPB	11 Sep 2017	13:13:56	U	SMPL
I5167-03 - 1	CPS083061	-271	-0.0766	-	7470A	PPB	11 Sep 2017	13:16:04	U	SMPL
I5167-04 - 1	CPS073061	-261	-0.0739	-	7470A	PPB	11 Sep 2017	13:18:12	U	SMPL
I5167-05 - 1	CPS073062	-269	-0.0761	-	7470A	PPB	11 Sep 2017	13:20:20	U	SMPL
I5167-06 - 1	CPS133061	-218	-0.0622	-	7470A	PPB	11 Sep 2017	13:22:29	U	SMPL
I5167-07 - 1	CPS043061	-207	-0.0592	-	7470A	PPB	11 Sep 2017	13:24:36	U	SMPL
CCV85 - 1	CCV85	18245	4.9663	-	7470A	PPB	11 Sep 2017	13:26:43	U	SMPL
CCB85 - 1	CCB85	-103	-0.0309	-	7470A	PPB	11 Sep 2017	13:28:50	U	SMPL
I5167-08 - 1	CPS043065	-274	-0.0774	-	7470A	PPB	11 Sep 2017	13:30:57	U	SMPL
I5167-09 - 1	CPS113061	-202	-0.0578	-	7470A	PPB	11 Sep 2017	13:33:05	U	SMPL
I5167-11 - 1	CPS123061	-232	-0.0660	-	7470A	PPB	11 Sep 2017	13:37:19	U	SMPL
I5167-12 - 1	CPS103061	-362	-0.1014	-	7470A	PPB	11 Sep 2017	13:39:26	U	SMPL
I5167-13 - 1	CPS083061	-254	-0.0720	-	7470A	PPB	11 Sep 2017	13:41:34	U	SMPL
I5167-14 - 1	CPS073061	-226	-0.0644	-	7470A	PPB	11 Sep 2017	13:43:42	U	SMPL
I5167-15 - 1	CPS073062	-255	-0.0722	-	7470A	PPB	11 Sep 2017	13:45:50	U	SMPL
I5167-16 - 1	CPS133061	-239	-0.0679	-	7470A	PPB	11 Sep 2017	13:47:58	U	SMPL
I5167-17 - 1	CPS043061	-270	-0.0763	-	7470A	PPB	11 Sep 2017	13:50:05	U	SMPL
CCV86 - 1	CCV86	17996	4.8985	-	7470A	PPB	11 Sep 2017	13:52:12	U	SMPL
CCB86 - 1	CCB86	-122	-0.0360	-	7470A	PPB	11 Sep 2017	13:54:20	U	SMPL
I5167-18 - 1	CPS113061	-286	-0.0807	-	7470A	PPB	11 Sep 2017	13:56:27	U	SMPL
I5174-01 - 1	20170612-FROMWIER	-186	-0.0535	-	7470A	PPB	11 Sep 2017	14:00:41	U	SMPL
I5158-03LX5 - 1		-14	-0.0066	-	7470A	PPB	11 Sep 2017	14:02:48	U	SMPL
I5158-03A - 1		14299	3.8916	-	7470A	PPB	11 Sep 2017	14:04:56	U	SMPL
I5164-01LX5 - 1		-179	-0.0516	-	7470A	PPB	11 Sep 2017	14:07:03	U	SMPL
I5164-01A - 1		14104	3.8385	-	7470A	PPB	11 Sep 2017	14:09:11	U	SMPL
CCV87 - 1	CCV87	17964	4.8898	-	7470A	PPB	11 Sep 2017	14:11:20	U	SMPL
CCB87 - 1	CCB87	-134	-0.0393	-	7470A	PPB	11 Sep 2017	14:13:29	U	SMPL