


ISM02.4

LB89994
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4038E



A= 0.0000e+000

B= 2.5174e-004

C= -1.2800e-001

Rho= 0.9998885

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%rD
0.00	0.000	-0.025	-0.025	409	0.000	409					-11
0.2	0.200	0.178	-0.022	1217	0.0 %	1217					4
2.5	2.500	2.591	0.091	10800	0.0 %	10800					0
5.0	5.000	5.003	0.003	20381	0.0 %	20381					-1
7.5	7.500	7.414	-0.086	29960	0.0 %	29960					0
10.0	10.000	10.039	0.039	40386	0.0 %	40386					

nb
9/11/12

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	409	-		0 ISM02.4	PPB	9/11/2017 15:38	Std
	0.2 S0.2	1217	-		0.2 ISM02.4	PPB	9/11/2017 15:40	Std
	2.5 S2.5	10800	-		2.5 ISM02.4	PPB	9/11/2017 15:43	Std
	5 S5	20381	-		5 ISM02.4	PPB	9/11/2017 15:45	Std
	7.5 S7.5	29960	-		7.5 ISM02.4	PPB	9/11/2017 15:47	Std
	10 S10	40386	-		10 ISM02.4	PPB	9/11/2017 15:50	Std
ICV	ICV	15630	3.8067 -		ISM02.4	PPB	9/11/2017 15:53	SMPL
ICB	ICB	219	-0.0729 -		ISM02.4	PPB	9/11/2017 15:55	SMPL
CCV088	CCV088	20547	5.0446 -		ISM02.4	PPB	9/11/2017 15:57	SMPL
CCB088	CCB088	222	-0.0721 -		ISM02.4	PPB	9/11/2017 15:59	SMPL
PB102244BL	PBW003	373	-0.0341 -		ISM02.4	PPB	9/11/2017 16:04	SMPL
I5192-01	MBDD12	359	-0.0376 -		ISM02.4	PPB	9/11/2017 16:11	SMPL
I5192-02	MBDD14	370	-0.0349 -		ISM02.4	PPB	9/11/2017 16:16	SMPL
I5192-03	MBDD14D	339	-0.0427 -		ISM02.4	PPB	9/11/2017 16:18	SMPL
I5192-04	MBDD14S	4003	0.8797 -		ISM02.4	PPB	9/11/2017 16:21	SMPL
PB102245BL	PBW003	335	-0.0437 -		ISM02.4	PPB	9/11/2017 16:23	SMPL
I5193-01	MBDD11	351	-0.0396 -		ISM02.4	PPB	9/11/2017 16:25	SMPL
I5193-02	MBDD13	341	-0.0422 -		ISM02.4	PPB	9/11/2017 16:28	SMPL
I5193-03	MBDD13D	366	-0.0359 -		ISM02.4	PPB	9/11/2017 16:30	SMPL
I5193-04	MBDD13S	3966	0.8704 -		ISM02.4	PPB	9/11/2017 16:32	SMPL
CCV089	CCV089	20596	5.0569 -		ISM02.4	PPB	9/11/2017 16:34	SMPL
CCB089	CCB089	90	-0.1053 -		ISM02.4	PPB	9/11/2017 16:37	SMPL