



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89999

Temp.: 20.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/11/2017	16:15
2	CAL2	1	Water			7.00	09/11/2017	16:16
3	CAL3	1	Water			10.02	09/11/2017	16:17
4	ICV	1	Water			7.00	09/11/2017	16:18
5	CCV1	1	Water			2.02	09/11/2017	16:20
6	I5202-01	1	Water			8.03	09/11/2017	16:22
7	I5202-01DUP	1	Water			8.04	09/11/2017	16:23
8	CCV2	1	Water			12.01	09/11/2017	16:25

WORKLIST(Hardcopy Internal Chain)

2B89999

WorkList Name : PH I5202

WorkList ID : 102787

Date : 9/11/2017 4:07:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/11/2017	Water	I5202-01-17	pH	Cool 4 deg C	PSEG01	D52	A5972-OT	09/11/2017	9045C

Date/Time 09-11-17 4:07 PM
 Received by: JD
 Relinquished by: CP

Date/Time 09-11-17 5:00 PM
 Received by: CP
 Relinquished by: JD