



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB90004

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	09/11/2017	16:30
2	CAL2	1	Water			7.01	09/11/2017	16:31
3	CAL3	1	Water			10.03	09/11/2017	16:33
4	ICV	1	Water			6.99	09/11/2017	16:35
5	CCV1	1	Water			2.01	09/11/2017	16:36
6	I5201-03	1	Solid	20.02	20	6.81	09/11/2017	16:37
7	I5203-02	1	Solid	20.03	20	8.78	09/11/2017	16:38
8	I5208-03	1	Solid	20.01	20	7.30	09/11/2017	16:40
9	I5208-06	1	Solid	20.03	20	7.18	09/11/2017	16:41
10	I5208-09	1	Solid	20.02	20	8.52	09/11/2017	16:42
11	I5208-12	1	Solid	20.04	20	9.25	09/11/2017	16:44
12	I5208-15	1	Solid	20.02	20	9.31	09/11/2017	16:45
13	I5209-03	1	Solid	20.03	20	8.80	09/11/2017	16:46
14	I5209-03DUP	1	Solid	20.05	20	8.81	09/11/2017	16:47
15	CCV2	1	Water			12.01	09/11/2017	16:49

2B90004

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I5203 WorkList ID : 102786 Date : 9/11/2017 4:07:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/14/2017	Solid	I5201-03	Corrosivity	Cool 4 deg C	PSEG05	D52	SU-03-091117	09/11/2017	9045C
09/14/2017	Solid	I5203-02	Corrosivity	Cool 4 deg C	PSEG01	D52	A4129	09/11/2017	9045C
	Solid	I5208-03	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-091117-A	09/11/2017	9045C
	Solid	I5208-06	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-091117-B	09/11/2017	9045C
	Solid	I5208-09	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-091117-C	09/11/2017	9045C
	Solid	I5208-12	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-091117-D	09/11/2017	9045C
	Solid	I5208-15	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-091117-E	09/11/2017	9045C
	Solid	I5209-03	Corrosivity	Cool 4 deg C	PSEG05	D52	OR-03-91117	09/11/2017	9045C

Date/Time 09-11-17 4:07 PM
 Received by: JD
 Relinquished by: CP

Date/Time 09-11-17 6:30 PM
 Received by: CP
 Relinquished by: JD