


7470A

LB90011
INSTRUMENT ID: CV 1Linear  μ Abs.:

35721



A= 0.0000e+000

B= 2.7997e-004

C= 5.2547e-002

Rho= 0.9999169

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	-81	0.000	-81				
0.2	0.200	0.193	-0.007	502	0.0 %	502				
2.5	2.500	2.477	-0.023	8659	0.0 %	8659				
5.0	5.000	5.036	0.036	17799	0.0 %	17799				
7.5	7.500	7.411	-0.089	26282	0.0 %	26282				
10.0	10.000	10.054	0.054	35721	0.0 %	35721				

LB90011 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 12 Sep 2017 09:51:20					
Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-81	-	0.0000	7470A	PPB	12 Sep 2017 10:01:59	S	Std
0.2 - 1	50.2	502	-	0.2000	7470A	PPB	12 Sep 2017 10:04:08	S	Std
2.5 - 1	52.5	8659	-	2.5000	7470A	PPB	12 Sep 2017 10:06:16	S	Std
5.0 - 1	55	17799	-	5.0000	7470A	PPB	12 Sep 2017 10:08:25	S	Std
7.5 - 1	57.5	26282	-	7.5000	7470A	PPB	12 Sep 2017 10:10:32	S	Std
10.0 - 1	510	35721	-	10.0000	7470A	PPB	12 Sep 2017 10:12:39	S	Std
ICV51 - 1	ICV51	13982	3.9672	-	7470A	PPB	12 Sep 2017 10:15:15	U	SMPL
ICB51 - 1	ICB51	-188	-0.0001	-	7470A	PPB	12 Sep 2017 10:17:21	U	SMPL
CCV90 - 1	CCV90	17838	5.0467	-	7470A	PPB	12 Sep 2017 10:19:28	U	SMPL
CCB90 - 1	CCB90	-248	-0.0169	-	7470A	PPB	12 Sep 2017 10:21:34	U	SMPL
CRA - 1	CRA	520	0.1981	-	7470A	PPB	12 Sep 2017 10:23:41	U	SMPL
HighStd - 1	HighStd	35660	10.0364	-	7470A	PPB	12 Sep 2017 10:25:48	U	SMPL
ChkStd - 1	ChkStd	25503	7.1927	-	7470A	PPB	12 Sep 2017 10:27:54	U	SMPL
PB102243BL - 1	PBW	-277	-0.0250	-	7470A	PPB	12 Sep 2017 10:30:03	U	SMPL
PB102243BS - 1	LCSW	13458	3.8204	-	7470A	PPB	12 Sep 2017 10:32:12	U	SMPL
PB102243TB - 1	PB102173TB	-175	0.0036	-	7470A	PPB	12 Sep 2017 10:34:20	U	SMPL
I5171-07 - 1	RT-2342-2669-3025-3091-3873	-122	0.0184	-	7470A	PPB	12 Sep 2017 10:36:29	U	SMPL
I5172-03 - 1	S32-9020	-122	0.0184	-	7470A	PPB	12 Sep 2017 10:38:37	U	SMPL
I5172-06 - 1	S32-9021	-143	0.0125	-	7470A	PPB	12 Sep 2017 10:40:46	U	SMPL
I5175-03 - 1	OILY-DRUM-COMP	-148	0.0111	-	7470A	PPB	12 Sep 2017 10:42:52	U	SMPL
CCV91 - 1	CCV91	17895	5.0627	-	7470A	PPB	12 Sep 2017 10:44:59	U	SMPL
CCB91 - 1	CCB91	-204	-0.0046	-	7470A	PPB	12 Sep 2017 10:47:06	U	SMPL
I5175-06 - 1	NORTH-EAST-SP-COMP	-109	0.0220	-	7470A	PPB	12 Sep 2017 10:49:13	U	SMPL
I5183-03 - 1	SASP2P-TP-BM-1(0-5)COMP	-110	0.0217	-	7470A	PPB	12 Sep 2017 10:51:20	U	SMPL
I5184-03 - 1	OK-02-090817	-109	0.0220	-	7470A	PPB	12 Sep 2017 10:53:26	U	SMPL
I5186-02 - 1	284-GRAB	-133	0.0153	-	7470A	PPB	12 Sep 2017 10:55:33	U	SMPL
I5186-02DUP - 1	284-GRABDUP	-118	0.0195	-	7470A	PPB	12 Sep 2017 10:57:40	U	SMPL
I5186-02MS - 1	284-GRABMS	12243	3.4803	-	7470A	PPB	12 Sep 2017 10:59:48	U	SMPL
I5186-02MSD - 1	284-GRABMSD	12411	3.5273	-	7470A	PPB	12 Sep 2017 11:01:56	U	SMPL
I5186-02LX5 - 1		-202	-0.0040	-	7470A	PPB	12 Sep 2017 11:04:04	U	SMPL
I5186-02A - 1		14526	4.1195	-	7470A	PPB	12 Sep 2017 11:06:14	U	SMPL
CCV92 - 1	CCV92	17720	5.0137	-	7470A	PPB	12 Sep 2017 11:08:20	U	SMPL
CCB92 - 1	CCB92	-249	-0.0172	-	7470A	PPB	12 Sep 2017 11:10:28	U	SMPL