

245.1


LB 90015

INSTRUMENT ID: CY1

Linear

 μ Abs.:

33375



A= 0.0000e+000

B= 2.9857e-004

C= 3.9779e-002

Rho= 0.9999789

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	-150	0.000	-150				
0.2	0.200	0.184	-0.016	482	0.0 %	482				
2.5	2.500	2.548	0.048	8402	0.0 %	8402				
5.0	5.000	4.972	-0.028	16519	0.0 %	16519				
7.5	7.500	7.497	-0.003	24976	0.0 %	24976				
10.0	10.000	10.004	0.004	33375	0.0 %	33375				

LB90015 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Sep 2017 11:22:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-150	-	0.0000	245.1	PPB	12 Sep 2017 11:24:36	S	Std
0.2 - 1	SD	482	-	0.2000	245.1	PPB	12 Sep 2017 11:26:45	S	Std
2.5 - 1	SD	8402	-	2.5000	245.1	PPB	12 Sep 2017 11:28:53	S	Std
5.0 - 1	SD	16519	-	5.0000	245.1	PPB	12 Sep 2017 11:31:02	S	Std
7.5 - 1	SD	24976	-	7.5000	245.1	PPB	12 Sep 2017 11:33:09	S	Std
10.0 - 1	SD	33375	-	10.0000	245.1	PPB	12 Sep 2017 11:37:47	S	Std
ICV52 - 1	ICV52	12660	3.8196	-245.1		PPB	12 Sep 2017 11:40:15	U	SMPL
ICB52 - 1	ICB52	-306	-0.0516	-245.1		PPB	12 Sep 2017 11:42:22	U	SMPL
CCV93 - 1	CCV93	16376	4.9291	-245.1		PPB	12 Sep 2017 11:44:28	U	SMPL
CCB93 - 1	CCB93	-268	-0.0402	-245.1		PPB	12 Sep 2017 11:46:35	U	SMPL
CRA - 1	CRA	496	0.1879	-245.1		PPB	12 Sep 2017 11:48:43	U	SMPL
HighStd - 1	HighStd	33614	10.0758	-245.1		PPB	12 Sep 2017 11:50:49	U	SMPL
ChkStd - 1	ChkStd	22800	6.8471	-245.1		PPB	12 Sep 2017 11:52:56	U	SMPL
PB102252BL - 1	PBW	-316	-0.0546	-245.1		PPB	12 Sep 2017 11:55:06	U	SMPL
PB102252BS - 1	LCSW	13551	4.0856	-245.1		PPB	12 Sep 2017 11:57:14	U	SMPL
I5141-01 - 1	15220032-1	-363	-0.0686	-245.1		PPB	12 Sep 2017 11:59:21	U	SMPL
I5142-01 - 1	32220008-3	-293	-0.0477	-245.1		PPB	12 Sep 2017 12:01:30	U	SMPL
I5142-01DUP - 1	32220008-3DUP	-252	-0.0355	-245.1		PPB	12 Sep 2017 12:03:38	U	SMPL
I5142-01MS - 1	32220008-3MS	10856	3.2810	-245.1		PPB	12 Sep 2017 12:05:46	U	SMPL
I5142-01MSD - 1	32220008-3MSD	10361	3.1332	-245.1		PPB	12 Sep 2017 12:07:53	U	SMPL
CCV94 - 1	CCV94	16086	4.8425	-245.1		PPB	12 Sep 2017 12:10:01	U	SMPL
CCB94 - 1	CCB94	-318	-0.0552	-245.1		PPB	12 Sep 2017 12:12:08	U	SMPL
I5180-01 - 1	090717-TIDO-F-D-B	-247	-0.0340	-245.1		PPB	12 Sep 2017 12:14:16	U	SMPL
I5180-02 - 1	090717-TIDO-F-D-S	-251	-0.0352	-245.1		PPB	12 Sep 2017 12:16:22	U	SMPL
I5180-03 - 1	090717-TIDO-F-U-B	-224	-0.0271	-245.1		PPB	12 Sep 2017 12:18:29	U	SMPL
I5180-04 - 1	090717-TIDO-F-U-S	-264	-0.0390	-245.1		PPB	12 Sep 2017 12:20:36	U	SMPL
I5181-01 - 1	090817-TIDO-F-D-M	-258	-0.0373	-245.1		PPB	12 Sep 2017 12:22:44	U	SMPL
I5181-02 - 1	090817-TIDO-F-U-M	-221	-0.0262	-245.1		PPB	12 Sep 2017 12:24:52	U	SMPL
I5142-01LX5 - 1		-217	-0.0250	-245.1		PPB	12 Sep 2017 12:27:00	U	SMPL
I5142-01A - 1		11334	3.4237	-245.1		PPB	12 Sep 2017 12:29:08	U	SMPL
CCV95 - 1	CCV95	16360	4.9243	-245.1		PPB	12 Sep 2017 12:31:16	U	SMPL
CCB95 - 1	CCB95	-257	-0.0370	-245.1		PPB	12 Sep 2017 12:33:24	U	SMPL