



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90024

Temp.: 20.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/12/2017	14:30
2	CAL2	1	Water			7.01	09/12/2017	14:31
3	CAL3	1	Water			10.04	09/12/2017	14:33
4	ICV	1	Water			7.00	09/12/2017	14:35
5	CCV1	1	Water			2.01	09/12/2017	14:36
6	I5214-04	1	Water			5.69	09/12/2017	14:37
7	I5214-04DUP	1	Water			5.60	09/12/2017	14:38
8	CCV2	1	Water			12.01	09/12/2017	14:40

1590024

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5214

WorkList ID : 102826

Date : 9/12/2017 2:16:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5214-04 ✓	pH	Cool 4 deg C	MAJO01	D32	DRAIN-WATER-TANK-1	09/12/2017	SM 4500-PH B

Date/Time 09-12-17 2:16 PM
Received by: JV
Relinquished by: CF

Date/Time 09-12-17 3:00 PM
Received by: CF
Relinquished by: JD