

ISM02.4


LB 90027

INSTRUMENT ID: CVD

Linear

 μ Abs.:

41126



A= 0.0000e+000

B= 2.4541e-004

C= -1.0952e-001

Rho= 0.9999615

Accept=Accepted

slope
y = 1.01e3 x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.000	-0.000	445	0.000	445				
0.2	0.200	0.198	-0.002	1254	0.0 %	1254				
2.5	2.500	2.523	0.023	10727	0.0 %	10727				
5.0	5.000	4.946	-0.054	20600	0.0 %	20600				
7.5	7.500	7.550	0.050	31213	0.0 %	31213				
10.0	10.000	9.983	-0.017	41126	0.0 %	41126				

9/0D

-1
1
-1
1
0

NB
9/12/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	445	-		0 ISM02.4	PPB	9/12/2017 14:59	Std
	0.2 S0.2	1254	-		0.2 ISM02.4	PPB	9/12/2017 15:01	Std
	2.5 S2.5	10727	-		2.5 ISM02.4	PPB	9/12/2017 15:04	Std
	5 S5	20600	-		5 ISM02.4	PPB	9/12/2017 15:06	Std
	7.5 S7.5	31213	-		7.5 ISM02.4	PPB	9/12/2017 15:08	Std
	10 S10	41126	-		10 ISM02.4	PPB	9/12/2017 15:10	Std
ICV	ICV	15929	3.7995 -		ISM02.4	PPB	9/12/2017 15:13	SMPL
ICB	ICB	239	-0.0509 -		ISM02.4	PPB	9/12/2017 15:15	SMPL
CCV096	CCV096	20981	5.0393 -		ISM02.4	PPB	9/12/2017 15:18	SMPL
CCB096	CCB096	248	-0.0487 -		ISM02.4	PPB	9/12/2017 15:20	SMPL
PB102276BL	PBW011	448	0.0004 -		ISM02.4	PPB	9/12/2017 15:25	SMPL
I5192-05	MBDD16	465	0.0046 -		ISM02.4	PPB	9/12/2017 15:30	SMPL
PB102279BL	PBW011	424	-0.0055 -		ISM02.4	PPB	9/12/2017 15:32	SMPL
I5193-05	MBDD15	436	-0.0025 -		ISM02.4	PPB	9/12/2017 15:34	SMPL
CCV097	CCV097	21142	5.0788 -		ISM02.4	PPB	9/12/2017 15:43	SMPL
CCB097	CCB097	229	-0.0533 -		ISM02.4	PPB	9/12/2017 15:45	SMPL