



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB90035

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	09/12/2017	17:00
2	CAL2	1	Water			7.01	09/12/2017	17:01
3	CAL3	1	Water			10.03	09/12/2017	17:02
4	ICV	1	Water			7.00	09/12/2017	17:03
5	CCV1	1	Water			2.01	09/12/2017	17:05
6	I5219-01	1	Solid	20.02	20	7.28	09/12/2017	17:06
7	I5219-01DUP	1	Solid	20.04	20	7.29	09/12/2017	17:09
8	CCV2	1	Water			12.01	09/12/2017	17:10

LB90035

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I5219 WorkList ID : 102830 Date : 9/12/2017 4:44:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5219-01	pH	Cool 4 deg C	TULL02	E21	FK-PS-01-015	09/12/2017	9045C

Date/Time 09.12.17 14:44:12
Received by: JP
Relinquished by: JP

Date/Time 09.12.17 6:00 PM
Received by: CR
Relinquished by: JP